

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110922\
 Data File : VD074834.D
 Acq On : 09 Nov 2022 18:00
 Operator : VA/SY
 Sample : N5511-03
 Misc : 5.08G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 PL-04-110722

Quant Time: Nov 10 00:37:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 03 02:30:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	190393	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	309528	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	264443	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	98257	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	58852	39.028	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	78.060%
35) Dibromofluoromethane	7.810	113	87506	47.792	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	95.580%
50) Toluene-d8	10.269	98	239454	35.852	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	71.700%
62) 4-Bromofluorobenzene	12.575	95	69492	36.036	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	72.080%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110922\
Data File : VD074834.D
Acq On : 09 Nov 2022 18:00
Operator : VA/SY
Sample : N5511-03
Misc : 5.08G/5.00ml/MSVOA_D/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
PL-04-110722

Quant Time: Nov 10 00:37:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 03 02:30:45 2022
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.875 min Scan# 1069
Delta R.T. 0.000 min
Lab File: VD074834.D
Acq: 09 Nov 2022 18:00




Instrument :
MSVOA_D
ClientSampleId :
PL-04-110722

Tgt Ion:168 Resp: 190393
Ion Ratio Lower Upper
168 100
99 53.7 40.9 61.3



#33
1,2-Dichloroethane-d4
Concen: 39.028 ug/l
RT: 8.228 min Scan# 1129
Delta R.T. 0.000 min
Lab File: VD074834.D
Acq: 09 Nov 2022 18:00

Tgt Ion: 65 Resp: 58852
Ion Ratio Lower Upper
65 100
67 56.2 0.0 110.2





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.775 min Scan# 11

Delta R.T. 0.000 min

Lab File: VD074834.D

Acq: 09 Nov 2022 18:00

Instrument :

MSVOA_D

ClientSampleId :

PL-04-110722





Tgt Ion:114 Resp: 309528

Ion Ratio Lower Upper

114 100

63 16.9 0.0 33.2

88 13.8 0.0 28.2



#35

Dibromofluoromethane

Concen: 47.792 ug/l

RT: 7.810 min Scan# 1058

Delta R.T. 0.005 min

Lab File: VD074834.D

Acq: 09 Nov 2022 18:00




Tgt Ion:113 Resp: 87506


Ion Ratio Lower Upper

113 100

111 102.8 82.3 123.5

192 18.6 16.4 24.6









#50
Toluene-d8
Concen: 35.852 ug/l
RT: 10.269 min Scan# 1476
Delta R.T. 0.000 min
Lab File: VD074834.D
Acq: 09 Nov 2022 18:00




Instrument :
MSVOA_D
ClientSampleId :
PL-04-110722


Tgt Ion: 98 Resp: 239454
Ion Ratio Lower Upper
98 100
100 63.1 51.4 77.0



#62
4-Bromofluorobenzene
Concen: 36.036 ug/l
RT: 12.575 min Scan# 1868
Delta R.T. -0.000 min
Lab File: VD074834.D
Acq: 09 Nov 2022 18:00


Tgt Ion: 95 Resp: 69492
Ion Ratio Lower Upper
95 100
174 96.2 0.0 197.0
176 90.3 0.0 182.6








#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.581 min Scan# 11
Delta R.T. -0.000 min
Lab File: VD074834.D
Acq: 09 Nov 2022 18:00


Instrument :
MSVOA_D
ClientSampleId :
PL-04-110722



Tgt Ion:117 Resp: 264443
Ion Ratio Lower Upper
117 100
82 52.2 41.6 62.4
119 31.9 25.2 37.8



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.516 min Scan# 2028
Delta R.T. 0.000 min
Lab File: VD074834.D
Acq: 09 Nov 2022 18:00



Tgt Ion:152 Resp: 98257
Ion Ratio Lower Upper
152 100
115 55.0 26.8 80.4
150 157.0 0.0 348.2

