

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111524\
 Data File : VD080029.D
 Acq On : 15 Nov 2024 17:13
 Operator : RP/MD
 Sample : P4837-03
 Misc : 5.10G/5ml/MSVOA_D/SOIL/A
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 3TRK-STONE

Quant Time: Nov 15 18:35:35 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111524S.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 15 16:25:47 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	163020	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	285688	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.580	117	260364	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	111904	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.228	65	94789	56.474	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	112.940%
35) Dibromofluoromethane	7.804	113	102705	54.017	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.040%
50) Toluene-d8	10.269	98	337340	47.074	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	94.140%
62) 4-Bromofluorobenzene	12.569	95	100920	42.924	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	85.840%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111524\
Data File : VD080029.D
Acq On : 15 Nov 2024 17:13
Operator : RP/MD
Sample : P4837-03
Misc : 5.10G/5ml/MSVOA_D/SOIL/A
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
3TRK-STONE

Quant Time: Nov 15 18:35:35 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111524S.M
Quant Title : SW846 8260
QLast Update : Fri Nov 15 16:25:47 2024
Response via : Initial Calibration

