

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111524\
 Data File : VD080031.D
 Acq On : 15 Nov 2024 18:07
 Operator : RP/MD
 Sample : P4835-03
 Misc : 5.01G/5ml/MSVOA_D/SOIL/A
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 345-2-STONE

Quant Time: Nov 15 18:36:08 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111524S.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 15 16:25:47 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	140414	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	241463	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	223457	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	94895	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	92182	63.763	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	127.520%
35) Dibromofluoromethane	7.805	113	95659	59.526	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	119.060%
50) Toluene-d8	10.269	98	306336	50.577	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	101.160%
62) 4-Bromofluorobenzene	12.569	95	100303	50.475	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	100.960%
Target Compounds						
16) Acetone	4.022	43	2900	9.245	ug/l	Qvalue # 82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111524\
Data File : VD080031.D
Acq On : 15 Nov 2024 18:07
Operator : RP/MD
Sample : P4835-03
Misc : 5.01G/5ml/MSVOA_D/SOIL/A
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
345-2-STONE

Quant Time: Nov 15 18:36:08 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111524S.M
Quant Title : SW846 8260
QLast Update : Fri Nov 15 16:25:47 2024
Response via : Initial Calibration

