

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111524\
 Data File : VD080032.D
 Acq On : 15 Nov 2024 18:34
 Operator : RP/MD
 Sample : P4834-03
 Misc : 5.08G/5ml/MSVOA_D/SOIL/A
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SLP-1-STONE

Quant Time: Nov 15 18:36:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111524S.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 15 16:25:47 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	29948	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	64383	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	73124	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	33416	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	44304	143.684	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 287.360%#	
35) Dibromofluoromethane	7.810	113	35727	83.379	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 166.760%#	
50) Toluene-d8	10.269	98	80698	49.969	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 99.940%	
62) 4-Bromofluorobenzene	12.569	95	31730	59.884	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 119.760%	

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111524\
Data File : VD080032.D
Acq On : 15 Nov 2024 18:34
Operator : RP/MD
Sample : P4834-03
Misc : 5.08G/5ml/MSVOA_D/SOIL/A
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SLP-1-STONE

Quant Time: Nov 15 18:36:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111524S.M
Quant Title : SW846 8260
QLast Update : Fri Nov 15 16:25:47 2024
Response via : Initial Calibration

