

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111524\
 Data File : VD080034.D
 Acq On : 15 Nov 2024 19:28
 Operator : RP/MD
 Sample : P4833-03
 Misc : 2.88G/5ml/MSVOA_D/SOIL/A
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 MH-731-VOC

Quant Time: Nov 15 22:26:49 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111524S.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 15 16:25:47 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	58	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.787	114	404	50.000	ug/l	0.01
63) Chlorobenzene-d5	11.593	117	276	50.000	ug/l	# 0.01
72) 1,4-Dichlorobenzene-d4	0.000	152	0	0.000	ug/l	-13.52
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.204	65	57	95.451	ug/l	-0.02
Spiked Amount	50.000	Range	50 - 163	Recovery	=	190.900%#
35) Dibromofluoromethane	7.793	113	225	83.682	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	167.360%#
50) Toluene-d8	10.269	98	488	48.155	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	96.320%
62) 4-Bromofluorobenzene	0.000	95	0	0.000	ug/l	
Spiked Amount	50.000	Range	30 - 143	Recovery	=	0.000%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111524\
Data File : VD080034.D
Acq On : 15 Nov 2024 19:28
Operator : RP/MD
Sample : P4833-03
Misc : 2.88G/5ml/MSVOA_D/SOIL/A
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
MH-731-VOC

Quant Time: Nov 15 22:26:49 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111524S.M
Quant Title : SW846 8260
QLast Update : Fri Nov 15 16:25:47 2024
Response via : Initial Calibration

