

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111524\
 Data File : VD080036.D
 Acq On : 15 Nov 2024 20:22
 Operator : RP/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ml/MSVOA_D/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 15 22:28:25 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111524S.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 15 16:25:47 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	94	0.00
2 T	Dichlorodifluoromethane	0.401	0.399	0.5	95	0.00
3 P	Chloromethane	0.295	0.262	11.2	89	0.00
4 C	Vinyl Chloride	0.337	0.313	7.1#	91	0.00
5 T	Bromomethane	0.374	0.322	13.9	99	0.00
6 T	Chloroethane	0.241	0.211	12.4	95	0.00
7 T	Trichlorofluoromethane	0.827	0.828	-0.1	101	0.00
8 T	Diethyl Ether	0.242	0.276	-14.0	114	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.517	0.550	-6.4	107	0.00
10 T	Methyl Iodide	0.640	0.717	-12.0	97	0.00
11 T	Tert butyl alcohol	0.027	0.031	-14.8	127	0.00
12 CM	1,1-Dichloroethene	0.518	0.558	-7.7#	106	0.00
13 T	Acrolein	0.051	0.051	0.0	100	0.00
14 T	Allyl chloride	0.734	0.792	-7.9	116	0.00
15 T	Acrylonitrile	0.107	0.127	-18.7	119	0.00
16 T	Acetone	0.112	0.124	-10.7	110	0.00
17 T	Carbon Disulfide	1.668	1.828	-9.6	110	0.00
18 T	Methyl Acetate	0.248	0.301	-21.4	132	0.00
19 T	Methyl tert-butyl Ether	1.057	1.269	-20.1	109	0.00
20 T	Methylene Chloride	0.599	0.673	-12.4	113	0.00
21 T	trans-1,2-Dichloroethene	0.573	0.625	-9.1	101	0.00
22 T	Diisopropyl ether	1.487	1.743	-17.2	105	0.00
23 T	Vinyl Acetate	0.852	1.013	-18.9	104	0.00
24 P	1,1-Dichloroethane	0.989	1.056	-6.8	102	0.00
25 T	2-Butanone	0.146	0.144	1.4	88	0.00
26 T	2,2-Dichloropropane	0.860	0.832	3.3	93	0.00
27 T	cis-1,2-Dichloroethene	0.657	0.741	-12.8	106	0.00
28 T	Bromochloromethane	0.440	0.492	-11.8	113	0.00
29 T	Tetrahydrofuran	0.080	0.101	-26.3#	111	0.00
30 C	Chloroform	1.042	1.159	-11.2#	107	0.00
31 T	Cyclohexane	0.838	0.847	-1.1	100	0.00
32 T	1,1,1-Trichloroethane	0.892	0.937	-5.0	101	0.00
33 S	1,2-Dichloroethane-d4	0.515	0.568	-10.3	101	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	95	0.00
35 S	Dibromofluoromethane	0.333	0.358	-7.5	101	0.00
36 T	1,1-Dichloropropene	0.455	0.492	-8.1	103	0.00
37 T	Ethyl Acetate	0.196	0.206	-5.1	101	0.00
38 T	Carbon Tetrachloride	0.497	0.526	-5.8	102	0.00
39 T	Methylcyclohexane	0.542	0.578	-6.6	99	0.00
40 TM	Benzene	1.418	1.560	-10.0	103	0.00
41 T	Methacrylonitrile	0.106	0.122	-15.1	111	0.00
42 TM	1,2-Dichloroethane	0.367	0.417	-13.6	108	0.00
43 T	Isopropyl Acetate	0.355	0.414	-16.6	107	0.00
44 TM	Trichloroethene	0.367	0.392	-6.8	103	0.00
45 C	1,2-Dichloropropane	0.341	0.385	-12.9#	104	0.00
46 T	Dibromomethane	0.198	0.224	-13.1	109	0.00
47 T	Bromodichloromethane	0.496	0.551	-11.1	105	0.00
48 T	Methyl methacrylate	0.169	0.207	-22.5	112	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111524\
 Data File : VD080036.D
 Acq On : 15 Nov 2024 20:22
 Operator : RP/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ml/MSVOA_D/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 15 22:28:25 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111524S.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 15 16:25:47 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.003	-50.0#	118	0.00
50 S	Toluene-d8	1.254	1.379	-10.0	99	0.00
51 T	4-Methyl-2-Pentanone	0.180	0.221	-22.8	109	0.00
52 CM	Toluene	0.881	1.006	-14.2#	104	0.00
53 T	t-1,3-Dichloropropene	0.442	0.495	-12.0	103	0.00
54 T	cis-1,3-Dichloropropene	0.524	0.595	-13.5	105	0.00
55 T	1,1,2-Trichloroethane	0.262	0.291	-11.1	104	0.00
56 T	Ethyl methacrylate	0.305	0.381	-24.9	110	0.00
57 T	1,3-Dichloropropane	0.427	0.489	-14.5	106	0.00
58 T	2-Chloroethyl Vinyl ether	0.115	0.145	-26.1#	112	0.00
59 T	2-Hexanone	0.132	0.159	-20.5	103	0.00
60 T	Dibromochloromethane	0.338	0.398	-17.8	109	0.00
61 T	1,2-Dibromoethane	0.247	0.285	-15.4	108	0.00
62 S	4-Bromofluorobenzene	0.411	0.463	-12.7	101	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	96	0.00
64 T	Tetrachloroethene	0.356	0.387	-8.7	108	0.00
65 PM	Chlorobenzene	1.091	1.180	-8.2	105	0.00
66 T	1,1,1,2-Tetrachloroethane	0.380	0.408	-7.4	103	0.00
67 C	Ethyl Benzene	1.830	2.011	-9.9#	103	0.00
68 T	m/p-Xylenes	0.715	0.795	-11.2	105	0.00
69 T	o-Xylene	0.643	0.728	-13.2	104	0.00
70 T	Styrene	1.148	1.319	-14.9	104	0.00
71 P	Bromoform	0.224	0.253	-12.9	108	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
73 T	Isopropylbenzene	3.224	3.484	-8.1	105	0.00
74 T	N-amyl acetate	0.710	0.814	-14.6	107	0.00
75 P	1,1,2,2-Tetrachloroethane	0.620	0.681	-9.8	110	0.00
76 T	1,2,3-Trichloropropane	0.442	0.449	-1.6	108	0.00
77 T	Bromobenzene	0.838	0.914	-9.1	108	0.00
78 T	n-propylbenzene	3.985	4.325	-8.5	105	0.00
79 T	2-Chlorotoluene	2.337	2.520	-7.8	105	0.00
80 T	1,3,5-Trimethylbenzene	2.729	3.020	-10.7	106	0.00
81 T	trans-1,4-Dichloro-2-butene	0.216	0.237	-9.7	106	0.00
82 T	4-Chlorotoluene	2.498	2.686	-7.5	105	0.00
83 T	tert-Butylbenzene	2.363	2.516	-6.5	103	0.00
84 T	1,2,4-Trimethylbenzene	2.741	3.075	-12.2	106	0.00
85 T	sec-Butylbenzene	3.494	3.797	-8.7	106	0.00
86 T	p-Isopropyltoluene	2.953	3.261	-10.4	106	0.00
87 T	1,3-Dichlorobenzene	1.697	1.814	-6.9	106	0.00
88 T	1,4-Dichlorobenzene	1.704	1.801	-5.7	107	0.00
89 T	n-Butylbenzene	2.778	3.013	-8.5	105	0.00
90 T	Hexachloroethane	0.633	0.654	-3.3	106	0.00
91 T	1,2-Dichlorobenzene	1.480	1.596	-7.8	107	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.094	0.099	-5.3	107	0.00
93 T	1,2,4-Trichlorobenzene	0.928	1.087	-17.1	116	0.00
94 T	Hexachlorobutadiene	0.501	0.590	-17.8	118	0.00
95 T	Naphthalene	1.576	1.940	-23.1	118	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111524\
Data File : VD080036.D
Acq On : 15 Nov 2024 20:22
Operator : RP/MD
Sample : VSTDCCC050
Misc : 5.00G/5ml/MSVOA_D/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Nov 15 22:28:25 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111524S.M
Quant Title : SW846 8260
QLast Update : Fri Nov 15 16:25:47 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.815	0.955	-17.2	114	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6