

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111924\
 Data File : VD080089.D
 Acq On : 19 Nov 2024 17:52
 Operator : RP/MD
 Sample : P4889-02
 Misc : 5.01G/5ml/MSVOA_D/SOIL/A
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 COMP-2

Quant Time: Nov 20 03:57:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111524S.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 15 16:25:47 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	158381	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	288332	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	256039	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	102603	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	89596	54.944	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	109.880%
35) Dibromofluoromethane	7.805	113	93023	48.476	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.960%
50) Toluene-d8	10.269	98	288498	39.889	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	79.780%
62) 4-Bromofluorobenzene	12.569	95	80352	33.862	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	67.720%
Target Compounds						
81) trans-1,4-Dichloro-2-b...	12.569	75	39344	88.877	ug/l	Qvalue # 15

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111924\
Data File : VD080089.D
Acq On : 19 Nov 2024 17:52
Operator : RP/MD
Sample : P4889-02
Misc : 5.01G/5ml/MSVOA_D/SOIL/A
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
COMP-2

Quant Time: Nov 20 03:57:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111524S.M
Quant Title : SW846 8260
QLast Update : Fri Nov 15 16:25:47 2024
Response via : Initial Calibration

