

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111924\
 Data File : VD080094.D
 Acq On : 19 Nov 2024 20:07
 Operator : RP/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 20 03:59:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111524S.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 15 16:25:47 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	81	0.00
2 T	Dichlorodifluoromethane	0.401	0.409	-2.0	83	0.00
3 P	Chloromethane	0.295	0.463	-56.9#	136	0.00
4 C	Vinyl Chloride	0.337	0.588	-74.5#	148	0.00
5 T	Bromomethane	0.374	0.507	-35.6#	134	0.00
6 T	Chloroethane	0.241	0.395	-63.9#	154#	0.00
7 T	Trichlorofluoromethane	0.827	0.856	-3.5	90	0.00
8 T	Diethyl Ether	0.242	0.282	-16.5	101	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.517	0.518	-0.2	87	0.00
10 T	Methyl Iodide	0.640	0.709	-10.8	82	0.00
11 T	Tert butyl alcohol	0.027	0.032	-18.5	114	0.00
12 CM	1,1-Dichloroethene	0.518	0.533	-2.9#	88	0.00
13 T	Acrolein	0.051	0.043	15.7	73	0.00
14 T	Allyl chloride	0.734	0.745	-1.5	94	0.00
15 T	Acrylonitrile	0.107	0.123	-15.0	99	0.00
16 T	Acetone	0.112	0.108	3.6	83	0.00
17 T	Carbon Disulfide	1.668	1.722	-3.2	89	0.00
18 T	Methyl Acetate	0.248	0.299	-20.6	113	0.00
19 T	Methyl tert-butyl Ether	1.057	1.268	-20.0	94	0.00
20 T	Methylene Chloride	0.599	0.648	-8.2	94	0.00
21 T	trans-1,2-Dichloroethene	0.573	0.611	-6.6	85	0.00
22 T	Diisopropyl ether	1.487	1.708	-14.9	89	0.00
23 T	Vinyl Acetate	0.852	0.976	-14.6	86	0.00
24 P	1,1-Dichloroethane	0.989	1.071	-8.3	89	0.00
25 T	2-Butanone	0.146	0.154	-5.5	81	0.00
26 T	2,2-Dichloropropane	0.860	0.838	2.6	81	0.00
27 T	cis-1,2-Dichloroethene	0.657	0.721	-9.7	89	0.00
28 T	Bromochloromethane	0.440	0.489	-11.1	97	0.00
29 T	Tetrahydrofuran	0.080	0.097	-21.3	93	0.00
30 C	Chloroform	1.042	1.158	-11.1#	92	0.00
31 T	Cyclohexane	0.838	0.770	8.1	78	0.00
32 T	1,1,1-Trichloroethane	0.892	0.946	-6.1	88	0.00
33 S	1,2-Dichloroethane-d4	0.515	0.578	-12.2	89	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	85	0.00
35 S	Dibromofluoromethane	0.333	0.354	-6.3	89	0.00
36 T	1,1-Dichloropropene	0.455	0.464	-2.0	86	0.00
37 T	Ethyl Acetate	0.196	0.216	-10.2	94	0.00
38 T	Carbon Tetrachloride	0.497	0.488	1.8	84	0.00
39 T	Methylcyclohexane	0.542	0.521	3.9	80	0.00
40 TM	Benzene	1.418	1.499	-5.7	89	0.00
41 T	Methacrylonitrile	0.106	0.098	7.5	80	0.00
42 TM	1,2-Dichloroethane	0.367	0.403	-9.8	93	0.00
43 T	Isopropyl Acetate	0.355	0.394	-11.0	91	0.00
44 TM	Trichloroethene	0.367	0.377	-2.7	89	0.00
45 C	1,2-Dichloropropane	0.341	0.373	-9.4#	90	0.00
46 T	Dibromomethane	0.198	0.220	-11.1	96	0.00
47 T	Bromodichloromethane	0.496	0.537	-8.3	91	0.00
48 T	Methyl methacrylate	0.169	0.196	-16.0	95	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111924\
 Data File : VD080094.D
 Acq On : 19 Nov 2024 20:07
 Operator : RP/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 20 03:59:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111524S.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 15 16:25:47 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	97	0.00
50 S	Toluene-d8	1.254	1.354	-8.0	86	0.00
51 T	4-Methyl-2-Pentanone	0.180	0.210	-16.7	93	0.00
52 CM	Toluene	0.881	0.970	-10.1#	90	0.00
53 T	t-1,3-Dichloropropene	0.442	0.485	-9.7	90	0.00
54 T	cis-1,3-Dichloropropene	0.524	0.571	-9.0	90	0.00
55 T	1,1,2-Trichloroethane	0.262	0.296	-13.0	94	0.00
56 T	Ethyl methacrylate	0.305	0.363	-19.0	93	0.00
57 T	1,3-Dichloropropane	0.427	0.481	-12.6	93	0.00
58 T	2-Chloroethyl Vinyl ether	0.115	0.126	-9.6	87	0.00
59 T	2-Hexanone	0.132	0.152	-15.2	88	0.00
60 T	Dibromochloromethane	0.338	0.384	-13.6	94	0.00
61 T	1,2-Dibromoethane	0.247	0.277	-12.1	94	0.00
62 S	4-Bromofluorobenzene	0.411	0.451	-9.7	88	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	87	0.00
64 T	Tetrachloroethene	0.356	0.351	1.4	90	0.00
65 PM	Chlorobenzene	1.091	1.120	-2.7	91	0.00
66 T	1,1,1,2-Tetrachloroethane	0.380	0.396	-4.2	91	0.00
67 C	Ethyl Benzene	1.830	1.905	-4.1#	89	0.00
68 T	m/p-Xylenes	0.715	0.751	-5.0	90	0.00
69 T	o-Xylene	0.643	0.698	-8.6	91	0.00
70 T	Styrene	1.148	1.276	-11.1	92	0.00
71 P	Bromoform	0.224	0.244	-8.9	95	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	0.00
73 T	Isopropylbenzene	3.224	3.264	-1.2	91	0.00
74 T	N-amyl acetate	0.710	0.753	-6.1	91	0.00
75 P	1,1,2,2-Tetrachloroethane	0.620	0.649	-4.7	97	0.00
76 T	1,2,3-Trichloropropane	0.442	0.477	-7.9	106	0.00
77 T	Bromobenzene	0.838	0.872	-4.1	95	0.00
78 T	n-propylbenzene	3.985	4.060	-1.9	91	0.00
79 T	2-Chlorotoluene	2.337	2.420	-3.6	93	0.00
80 T	1,3,5-Trimethylbenzene	2.729	2.868	-5.1	92	0.00
81 T	trans-1,4-Dichloro-2-butene	0.216	0.232	-7.4	96	0.00
82 T	4-Chlorotoluene	2.498	2.535	-1.5	92	0.00
83 T	tert-Butylbenzene	2.363	2.328	1.5	88	0.00
84 T	1,2,4-Trimethylbenzene	2.741	2.922	-6.6	93	0.00
85 T	sec-Butylbenzene	3.494	3.563	-2.0	91	0.00
86 T	p-Isopropyltoluene	2.953	3.048	-3.2	91	0.00
87 T	1,3-Dichlorobenzene	1.697	1.729	-1.9	93	0.00
88 T	1,4-Dichlorobenzene	1.704	1.714	-0.6	94	0.00
89 T	n-Butylbenzene	2.778	2.792	-0.5	90	0.00
90 T	Hexachloroethane	0.633	0.622	1.7	93	0.00
91 T	1,2-Dichlorobenzene	1.480	1.549	-4.7	96	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.094	0.099	-5.3	98	0.00
93 T	1,2,4-Trichlorobenzene	0.928	0.966	-4.1	95	0.00
94 T	Hexachlorobutadiene	0.501	0.499	0.4	92	0.00
95 T	Naphthalene	1.576	1.788	-13.5	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111924\
Data File : VD080094.D
Acq On : 19 Nov 2024 20:07
Operator : RP/MD
Sample : VSTDCCC050
Misc : 5.00G/5ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Nov 20 03:59:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111524S.M
Quant Title : SW846 8260
QLast Update : Fri Nov 15 16:25:47 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.815	0.894	-9.7	98	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6