

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111925\
 Data File : VD080541.D
 Acq On : 19 Nov 2025 09:50
 Operator : RP/MD
 Sample : VD1119SBL01
 Misc : 5.00G/5ml/MSVOA_D/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD1119SBL01

Quant Time: Nov 20 04:03:49 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111725S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 18 00:45:54 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.876	168	256365	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.776	114	420910	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.582	117	354162	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.517	152	153283	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.229	65	104377	48.490	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	96.980%
35) Dibromofluoromethane	7.806	113	120335	48.561	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.120%
50) Toluene-d8	10.270	98	500623	49.940	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.880%
62) 4-Bromofluorobenzene	12.570	95	136190	47.257	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	94.520%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111925\
Data File : VD080541.D
Acq On : 19 Nov 2025 09:50
Operator : RP/MD
Sample : VD1119SBL01
Misc : 5.00G/5ml/MSVOA_D/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD1119SBL01

Quant Time: Nov 20 04:03:49 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111725S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 18 00:45:54 2025
Response via : Initial Calibration

