

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111925\
 Data File : VD080546.D
 Acq On : 19 Nov 2025 13:01
 Operator : RP/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 20 04:06:09 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111725S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 18 00:45:54 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	86	0.00
2 T	Dichlorodifluoromethane	0.373	0.347	7.0	83	0.00
3 P	Chloromethane	0.602	0.536	11.0	78	0.00
4 C	Vinyl Chloride	0.713	0.684	4.1#	84	0.00
5 T	Bromomethane	0.529	0.490	7.4	82	0.00
6 T	Chloroethane	0.489	0.481	1.6	86	0.00
7 T	Trichlorofluoromethane	0.811	0.799	1.5	86	0.00
8 T	Diethyl Ether	0.224	0.176	21.4	68	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.513	0.508	1.0	87	0.00
10 T	Methyl Iodide	0.659	0.668	-1.4	79	0.00
11 T	Tert butyl alcohol	0.020	0.015	25.0#	63	-0.02
12 CM	1,1-Dichloroethene	0.540	0.501	7.2#	80	0.00
13 T	Acrolein	0.034	0.021	38.2#	61	0.00
14 T	Allyl chloride	0.843	0.695	17.6	71	0.00
15 T	Acrylonitrile	0.090	0.071	21.1	65	0.00
16 T	Acetone	0.061	0.046	24.6	64	0.00
17 T	Carbon Disulfide	1.736	1.555	10.4	78	0.00
18 T	Methyl Acetate	0.189	0.143	24.3	63	0.00
19 T	Methyl tert-butyl Ether	0.956	0.792	17.2	69	0.00
20 T	Methylene Chloride	0.620	0.530	14.5	74	0.00
21 T	trans-1,2-Dichloroethene	0.602	0.557	7.5	80	0.00
22 T	Diisopropyl ether	1.522	1.235	18.9	68	0.00
23 T	Vinyl Acetate	0.793	0.632	20.3	66	0.00
24 P	1,1-Dichloroethane	1.041	0.926	11.0	76	0.00
25 T	2-Butanone	0.096	0.071	26.0#	61	0.00
26 T	2,2-Dichloropropane	0.895	0.839	6.3	81	0.00
27 T	cis-1,2-Dichloroethene	0.701	0.624	11.0	77	0.00
28 T	Bromochloromethane	0.381	0.317	16.8	72	0.00
29 T	Tetrahydrofuran	0.064	0.047	26.6#	60	0.00
30 C	Chloroform	1.021	0.938	8.1#	78	0.00
31 T	Cyclohexane	0.997	0.873	12.4	80	0.00
32 T	1,1,1-Trichloroethane	0.897	0.867	3.3	83	0.00
33 S	1,2-Dichloroethane-d4	0.420	0.368	12.4	77	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	80	0.00
35 S	Dibromofluoromethane	0.294	0.285	3.1	80	0.00
36 T	1,1-Dichloropropene	0.455	0.461	-1.3	81	0.00
37 T	Ethyl Acetate	0.138	0.110	20.3	62	0.00
38 T	Carbon Tetrachloride	0.457	0.488	-6.8	85	0.00
39 T	Methylcyclohexane	0.600	0.615	-2.5	82	0.00
40 TM	Benzene	1.409	1.355	3.8	76	0.00
41 T	Methacrylonitrile	0.090	0.072	20.0	67	0.00
42 TM	1,2-Dichloroethane	0.294	0.273	7.1	73	0.00
43 T	Isopropyl Acetate	0.270	0.219	18.9	62	0.00
44 TM	Trichloroethene	0.356	0.358	-0.6	80	0.00
45 C	1,2-Dichloropropane	0.323	0.296	8.4#	73	0.00
46 T	Dibromomethane	0.159	0.146	8.2	72	0.00
47 T	Bromodichloromethane	0.430	0.414	3.7	74	0.00
48 T	Methyl methacrylate	0.134	0.112	16.4	65	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111925\
 Data File : VD080546.D
 Acq On : 19 Nov 2025 13:01
 Operator : RP/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 20 04:06:09 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111725S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 18 00:45:54 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.001	0.001	0.0	68	0.00
50 S	Toluene-d8	1.191	1.171	1.7	81	0.00
51 T	4-Methyl-2-Pentanone	0.125	0.102	18.4	62	0.00
52 CM	Toluene	0.891	0.870	2.4#	77	0.00
53 T	t-1,3-Dichloropropene	0.381	0.348	8.7	71	0.00
54 T	cis-1,3-Dichloropropene	0.491	0.457	6.9	73	0.00
55 T	1,1,2-Trichloroethane	0.204	0.182	10.8	70	0.00
56 T	Ethyl methacrylate	0.235	0.201	14.5	65	0.00
57 T	1,3-Dichloropropane	0.355	0.320	9.9	70	0.00
58 T	2-Chloroethyl Vinyl ether	0.132	0.122	7.6	70	0.00
59 T	2-Hexanone	0.088	0.070	20.5	62	0.00
60 T	Dibromochloromethane	0.270	0.248	8.1	71	0.00
61 T	1,2-Dibromoethane	0.196	0.178	9.2	70	0.00
62 S	4-Bromofluorobenzene	0.342	0.320	6.4	79	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	79	0.00
64 T	Tetrachloroethene	0.360	0.376	-4.4	83	0.00
65 PM	Chlorobenzene	1.113	1.107	0.5	77	0.00
66 T	1,1,1,2-Tetrachloroethane	0.348	0.353	-1.4	78	0.00
67 C	Ethyl Benzene	2.018	2.053	-1.7#	80	0.00
68 T	m/p-Xylenes	0.772	0.800	-3.6	81	0.00
69 T	o-Xylene	0.714	0.713	0.1	78	0.00
70 T	Styrene	1.178	1.175	0.3	77	0.00
71 P	Bromoform	0.163	0.157	3.7	72	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	79	0.00
73 T	Isopropylbenzene	4.244	4.388	-3.4	82	0.00
74 T	N-amyl acetate	0.658	0.518	21.3	59	0.00
75 P	1,1,2,2-Tetrachloroethane	0.532	0.458	13.9	65	0.00
76 T	1,2,3-Trichloropropane	0.389	0.416	-6.9	95	0.00
77 T	Bromobenzene	0.901	0.892	1.0	77	0.00
78 T	n-propylbenzene	5.092	5.213	-2.4	81	0.00
79 T	2-Chlorotoluene	2.801	2.770	1.1	78	0.00
80 T	1,3,5-Trimethylbenzene	3.355	3.408	-1.6	81	0.00
81 T	trans-1,4-Dichloro-2-butene	0.204	0.164	19.6	63	0.00
82 T	4-Chlorotoluene	2.848	2.809	1.4	79	0.00
83 T	tert-Butylbenzene	3.017	3.077	-2.0	80	0.00
84 T	1,2,4-Trimethylbenzene	3.301	3.345	-1.3	80	0.00
85 T	sec-Butylbenzene	4.466	4.645	-4.0	84	0.00
86 T	p-Isopropyltoluene	3.704	3.908	-5.5	84	0.00
87 T	1,3-Dichlorobenzene	1.790	1.839	-2.7	80	0.00
88 T	1,4-Dichlorobenzene	1.750	1.764	-0.8	81	0.00
89 T	n-Butylbenzene	3.532	3.649	-3.3	82	0.00
90 T	Hexachloroethane	0.723	0.728	-0.7	81	0.00
91 T	1,2-Dichlorobenzene	1.479	1.443	2.4	77	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.073	0.058	20.5	64	0.00
93 T	1,2,4-Trichlorobenzene	0.914	0.877	4.0	77	0.00
94 T	Hexachlorobutadiene	0.509	0.545	-7.1	87	0.00
95 T	Naphthalene	1.487	1.306	12.2	68	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111925\
Data File : VD080546.D
Acq On : 19 Nov 2025 13:01
Operator : RP/MD
Sample : VSTDCCC050
Misc : 5.00G/5ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Nov 20 04:06:09 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111725S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 18 00:45:54 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.740	0.697	5.8	76	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6