

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD120421\
 Data File : VD071153.D
 Acq On : 04 Dec 2021 19:40
 Operator : VA/SY
 Sample : M4927-09
 Misc : 5.62G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 PH1-BOT-10

Quant Time: Dec 06 04:33:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 29 17:54:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	284089	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	500184	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	532191	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	256105	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	150184	45.054	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	90.100%
35) Dibromofluoromethane	7.914	113	152505	47.880	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	95.760%
50) Toluene-d8	10.337	98	644308	53.156	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	106.320%
62) 4-Bromofluorobenzene	12.625	95	251680	59.901	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	119.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD120421\
Data File : VD071153.D
Acq On : 04 Dec 2021 19:40
Operator : VA/SY
Sample : M4927-09
Misc : 5.62G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
PH1-BOT-10

Quant Time: Dec 06 04:33:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 29 17:54:09 2021
Response via : Initial Calibration

