

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD120421\
 Data File : VD071156.D
 Acq On : 04 Dec 2021 21:04
 Operator : VA/SY
 Sample : M4927-12
 Misc : 5.92G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 PH1-BOT-13

Quant Time: Dec 06 04:34:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 29 17:54:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	6650	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	10820	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	13356	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	6322	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	5911	75.753	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	151.500%
35) Dibromofluoromethane	7.914	113	2741	39.782	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	79.560%
50) Toluene-d8	10.338	98	13490	51.449	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.900%
62) 4-Bromofluorobenzene	12.620	95	6059	66.664	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	133.320%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD120421\
Data File : VD071156.D
Acq On : 04 Dec 2021 21:04
Operator : VA/SY
Sample : M4927-12
Misc : 5.92G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
PH1-BOT-13

Quant Time: Dec 06 04:34:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 29 17:54:09 2021
Response via : Initial Calibration

