

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD120421\
 Data File : VD071142.D
 Acq On : 04 Dec 2021 14:33
 Operator : VA/SY
 Sample : VD1204SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD1204SBL01

Quant Time: Dec 06 03:42:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 29 17:54:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	238113	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	410895	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	376404	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	163661	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	149324	53.445	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	106.880%
35) Dibromofluoromethane	7.914	113	135467	51.773	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.540%
50) Toluene-d8	10.332	98	473615	47.565	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	95.120%
62) 4-Bromofluorobenzene	12.626	95	166459	48.227	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	96.460%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD120421\
Data File : VD071142.D
Acq On : 04 Dec 2021 14:33
Operator : VA/SY
Sample : VD1204SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD1204SBL01

Quant Time: Dec 06 03:42:43 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 29 17:54:09 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.979 min Scan# 1030
Delta R.T. -0.006 min
Lab File: VD071142.D
Acq: 04 Dec 2021 14:33




Instrument :
MSVOA_D
ClientSampleId :
VD1204SBL01


Tgt Ion:168 Resp: 238113
Ion Ratio Lower Upper
168 100
99 58.8 46.6 69.8



#33
1,2-Dichloroethane-d4
Concen: 53.445 ug/l
RT: 8.326 min Scan# 1089
Delta R.T. 0.000 min
Lab File: VD071142.D
Acq: 04 Dec 2021 14:33

Tgt Ion: 65 Resp: 149324
Ion Ratio Lower Upper
65 100
67 51.4 0.0 112.2









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.861 min Scan# 1180
Delta R.T. 0.000 min
Lab File: VD071142.D
Acq: 04 Dec 2021 14:33




Instrument :
MSVOA_D
ClientSampleId :
VD1204SBL01


Tgt Ion:114 Resp: 410895
Ion Ratio Lower Upper
114 100
63 22.4 0.0 41.8
88 15.7 0.0 32.8



#35
Dibromofluoromethane
Concen: 51.773 ug/l
RT: 7.914 min Scan# 1019
Delta R.T. -0.006 min
Lab File: VD071142.D
Acq: 04 Dec 2021 14:33

Tgt Ion:113 Resp: 135467
Ion Ratio Lower Upper
113 100
111 104.7 82.0 123.0
192 18.6 15.2 22.8









#50
Toluene-d8
Concen: 47.565 ug/l
RT: 10.332 min Scan# 1431
Delta R.T. -0.006 min
Lab File: VD071142.D
Acq: 04 Dec 2021 14:33




Instrument :
MSVOA_D
ClientSampleId :
VD1204SBL01


Tgt Ion: 98 Resp: 473615
Ion Ratio Lower Upper
98 100
100 65.8 51.8 77.6



#62
4-Bromofluorobenzene
Concen: 48.227 ug/l
RT: 12.626 min Scan# 1820
Delta R.T. -0.000 min
Lab File: VD071142.D
Acq: 04 Dec 2021 14:33

Tgt Ion: 95 Resp: 166459
Ion Ratio Lower Upper
95 100
174 75.2 0.0 159.4
176 72.7 0.0 154.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.638 min Scan# 117
Delta R.T. -0.000 min
Lab File: VD071142.D
Acq: 04 Dec 2021 14:33

Instrument :
MSVOA_D
ClientSampleId :
VD1204SBL01

Tgt Ion:117 Resp: 376404
Ion Ratio Lower Upper
117 100
82 55.5 44.7 67.1
119 30.8 26.2 39.4



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.567 min Scan# 1980
Delta R.T. 0.000 min
Lab File: VD071142.D
Acq: 04 Dec 2021 14:33

Tgt Ion:152 Resp: 163661
Ion Ratio Lower Upper
152 100
115 62.0 29.8 89.3
150 158.5 0.0 349.2

