

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD120621\
 Data File : VD071163.D
 Acq On : 06 Dec 2021 10:59
 Operator : VA/SY
 Sample : VD1206SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD1206SBL01

Quant Time: Dec 07 06:49:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 29 17:54:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	223286	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	386647	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	357110	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	157069	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	140199	53.511	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	107.020%
35) Dibromofluoromethane	7.914	113	129358	52.539	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	105.080%
50) Toluene-d8	10.332	98	455794	48.646	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	97.300%
62) 4-Bromofluorobenzene	12.626	95	159041	48.968	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	97.940%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD120621\
Data File : VD071163.D
Acq On : 06 Dec 2021 10:59
Operator : VA/SY
Sample : VD1206SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD1206SBL01

Quant Time: Dec 07 06:49:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 29 17:54:09 2021
Response via : Initial Calibration

