

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD120621\
 Data File : VD071169.D
 Acq On : 06 Dec 2021 15:56
 Operator : VA/SY
 Sample : M4954-01
 Misc : 5.08G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-1(2.5-3)

Quant Time: Dec 07 06:52:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 29 17:54:09 2021
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 12/07/2021
 Supervised By : Amit Patel 12/08/2021

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	280290	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	8.856	114	479420	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	402103	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	101749	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	145461	44.228	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	88.460%		
35) Dibromofluoromethane	7.909	113	153620	50.319	ug/l	-0.01
Spiked Amount 50.000	Range 54 - 147		Recovery =	100.640%		
50) Toluene-d8	10.332	98	580226	49.943	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery =	99.880%		
62) 4-Bromofluorobenzene	12.626	95	157973	39.227	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery =	78.460%		
Target Compounds						
16) Acetone	4.162	43	8246m	39.075	ug/l	Qvalue
20) Methylene Chloride	4.962	84	20465	2.171	ug/l #	87
64) Tetrachloroethene	10.873	164	4907	1.922	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD120621\
Data File : VD071169.D
Acq On : 06 Dec 2021 15:56
Operator : VA/SY
Sample : M4954-01
Misc : 5.08G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-1(2.5-3)

Quant Time: Dec 07 06:52:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 29 17:54:09 2021
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 12/07/2021
Supervised By :Amit Patel 12/08/2021

