

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121124\
 Data File : VD080133.D
 Acq On : 11 Dec 2024 17:57
 Operator : RP/MD
 Sample : P5232-08
 Misc : 6.72G/10ml/MSVOA_D/SOIL/A
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 E28S2

Quant Time: Dec 12 01:00:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM121124SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Dec 12 00:57:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.775	114	252580	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.581	117	186141	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.516	152	49222	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.270	65	71226	17.693	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	70.760%
7) Chloroethane-d5	2.793	69	66428	18.505	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	74.040%
11) 1,1-Dichloroethene-d2	3.916	65	24354	15.762	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	63.040%
21) 2-Butanone-d5	6.987	46	18825	20.944	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	41.880%
24) Chloroform-d	7.563	84	114768	16.457	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	65.840%
26) 1,2-Dichloroethane-d4	8.228	65	64119	17.374	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	69.480%#
32) Benzene-d6	8.199	84	201605	18.227	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	72.920%
36) 1,2-Dichloropropane-d6	9.210	67	78547	20.914	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	83.640%
41) Toluene-d8	10.269	98	150015	14.356	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	57.440%
43) trans-1,3-Dichloroprop...	10.522	79	22771	14.808	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	59.240%
47) 2-Hexanone-d5	10.875	63	12614	22.208	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	44.420%
56) 1,1,2,2-Tetrachloroeth...	12.651	84	50034	15.568	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	62.280%
66) 1,2-Dichlorobenzene-d4	13.810	152	28854	16.236	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	64.960%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121124\
Data File : VD080133.D
Acq On : 11 Dec 2024 17:57
Operator : RP/MD
Sample : P5232-08
Misc : 6.72G/10ml/MSVOA_D/SOIL/A
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
E28S2

Quant Time: Dec 12 01:00:05 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM121124SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Dec 12 00:57:48 2024
Response via : Initial Calibration

