

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121224\
 Data File : VD080152.D
 Acq On : 12 Dec 2024 18:59
 Operator : RP/MD
 Sample : P5155-10RE
 Misc : 4.86G/10ml/MSVOA_D/SOIL/B
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 E28T5RE

Quant Time: Dec 13 02:05:02 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM121124SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Dec 13 02:01:12 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.781	114	348232	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.581	117	301534	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.516	152	91573	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.270	65	251463	45.308	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	181.240%#
7) Chloroethane-d5	2.793	69	200766	40.566	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	162.280%#
11) 1,1-Dichloroethene-d2	3.911	65	47452	22.275	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	89.080%
21) 2-Butanone-d5	6.987	46	49798	40.185	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	80.360%
24) Chloroform-d	7.569	84	204410	21.260	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	85.040%
26) 1,2-Dichloroethane-d4	8.228	65	123922	24.356	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	97.440%
32) Benzene-d6	8.205	84	416033	23.219	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	92.880%
36) 1,2-Dichloropropane-d6	9.210	67	144255	23.711	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	94.840%
41) Toluene-d8	10.269	98	358694	21.190	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	84.760%
43) trans-1,3-Dichloroprop...	10.528	79	54524	21.889	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	87.560%
47) 2-Hexanone-d5	10.875	63	39573	43.009	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	86.020%
56) 1,1,2,2-Tetrachloroeth...	12.645	84	111722	21.459	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	85.840%
66) 1,2-Dichlorobenzene-d4	13.816	152	80260	24.275	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	97.120%
Target Compounds					Qvalue	
25) Chloroform	7.599	83	14266	1.241	ug/L	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121224\
Data File : VD080152.D
Acq On : 12 Dec 2024 18:59
Operator : RP/MD
Sample : P5155-10RE
Misc : 4.86G/10ml/MSVOA_D/SOIL/B
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
E28T5RE

Quant Time: Dec 13 02:05:02 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM121124SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Dec 13 02:01:12 2024
Response via : Initial Calibration

