

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD122718\
 Data File : VD060724.D
 Acq On : 27 Dec 2018 14:11
 Operator : VA/AP
 Sample : J6446-01RE
 Misc : 3.75g/5ml/MSVOA D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P001-SS022-0612-02RE

Quant Time: Dec 28 01:54:44 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D122018S.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 21 17:37:47 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	6.37	168	318360	50.00	ug/l	-0.04
34) 1,4-Difluorobenzene	7.41	114	503592	50.00	ug/l	-0.03
63) Chlorobenzene-d5	11.26	117	277526	50.00	ug/l	-0.02
72) 1,4-Dichlorobenzene-d4	13.35	152	68487	50.00	ug/l	-0.02
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.75	65	259053	117.75	ug/l	-0.04
Spiked Amount 50.000			Recovery	=	235.50%	
35) Dibromofluoromethane	6.29	113	271343	73.62	ug/l	-0.04
Spiked Amount 50.000			Recovery	=	147.24%	
50) Toluene-d8	9.42	98	431421	39.23	ug/l	-0.04
Spiked Amount 50.000			Recovery	=	78.46%	
62) 4-Bromofluorobenzene	12.41	95	87593	22.74	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	45.48%	
Target Compounds						
16) Acetone	3.13	43	17528	45.620	ug/l	99
20) Methylene Chloride	3.69	84	23771	5.957	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD122718\
Data File : VD060724.D
Acq On : 27 Dec 2018 14:11
Operator : VA/AP
Sample : J6446-01RE
Misc : 3.75g/5ml/MSVOA D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
P001-SS022-0612-02RE

Quant Time: Dec 28 01:54:44 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D122018S.M
Quant Title : SW846 8260
QLast Update : Fri Dec 21 17:37:47 2018
Response via : Initial Calibration

