

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD122718\
 Data File : VD060736.D
 Acq On : 27 Dec 2018 19:46
 Operator : VA/AP
 Sample : J6447-04RE
 Misc : 3.82g/5ml/MSVOA D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P001-SS024-1218-01RE

Quant Time: Dec 28 03:00:57 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D122018S.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 21 17:37:47 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	6.37	168	874802	50.00	ug/l	-0.04
34) 1,4-Difluorobenzene	7.41	114	1193073	50.00	ug/l	-0.04
63) Chlorobenzene-d5	11.25	117	830503	50.00	ug/l	-0.03
72) 1,4-Dichlorobenzene-d4	13.35	152	319451	50.00	ug/l	-0.03
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.75	65	381012	63.02	ug/l	-0.05
Spiked Amount 50.000			Recovery	=	126.04%	
35) Dibromofluoromethane	6.28	113	493443	56.51	ug/l	-0.05
Spiked Amount 50.000			Recovery	=	113.02%	
50) Toluene-d8	9.42	98	1140665	43.78	ug/l	-0.04
Spiked Amount 50.000			Recovery	=	87.56%	
62) 4-Bromofluorobenzene	12.40	95	303167	33.22	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	66.44%	
Target Compounds						
16) Acetone	3.13	43	375167	355.352	ug/l	Qvalue 100
20) Methylene Chloride	3.70	84	39599	3.612	ug/l	97
25) 2-Butanone	5.58	43	15903	7.872	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD122718\
Data File : VD060736.D
Acq On : 27 Dec 2018 19:46
Operator : VA/AP
Sample : J6447-04RE
Misc : 3.82g/5ml/MSVOA D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
P001-SS024-1218-01RE

Quant Time: Dec 28 03:00:57 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D122018S.M
Quant Title : SW846 8260
QLast Update : Fri Dec 21 17:37:47 2018
Response via : Initial Calibration

