

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122722\
 Data File : VD075055.D
 Acq On : 27 Dec 2022 16:15
 Operator : KP/SY
 Sample : N6183-03
 Misc : 5.03G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 R-3

Quant Time: Dec 28 01:53:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D12222S.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 23 05:54:19 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	7049	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	16725	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	15645	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	7618	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	5873	85.237	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 170.480%#	
35) Dibromofluoromethane	7.805	113	5192	52.755	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 105.520%	
50) Toluene-d8	10.269	98	21233	53.395	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 106.800%	
62) 4-Bromofluorobenzene	12.575	95	8120	63.031	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 126.060%	
Target Compounds						Qvalue
3) Chloromethane	2.146	50	616	7.000	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122722\
Data File : VD075055.D
Acq On : 27 Dec 2022 16:15
Operator : KP/SY
Sample : N6183-03
Misc : 5.03G/5.00ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
R-3

Quant Time: Dec 28 01:53:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122222S.M
Quant Title : SW846 8260
QLast Update : Fri Dec 23 05:54:19 2022
Response via : Initial Calibration

