



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122722\
 Data File : VD075049.D
 Acq On : 27 Dec 2022 09:50
 Operator : KP/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D12222S.M
 Title : SW846 8260
 Last Update : Fri Dec 23 05:54:19 2022



AutoFind: Scans 1867, 1868, 1869; Background Corrected with Scan 1859

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.5	16249	PASS
75	95	30	60	51.0	42605	PASS
95	95	100	100	100.0	83493	PASS
96	95	5	9	6.7	5609	PASS
173	174	0.00	2	1.1	701	PASS
174	95	50	100	76.4	63819	PASS
175	174	5	9	7.8	4953	PASS
176	174	95	101	97.4	62152	PASS
177	176	5	9	6.9	4275	PASS