



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122923\
Data File : VD078097.D
Acq On : 29 Dec 2023 08:39
Operator : RP/MD
Sample : BFB
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122623S.M
Title : SW846 8260
Last Update : Wed Dec 27 03:40:16 2023



AutoFind: Scans 1867, 1868, 1869; Background Corrected with Scan 1859

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.3	14113	PASS
75	95	30	60	48.1	39251	PASS
95	95	100	100	100.0	81573	PASS
96	95	5	9	6.2	5073	PASS
173	174	0.00	2	1.1	860	PASS
174	95	50	100	94.7	77251	PASS
175	174	5	9	7.9	6113	PASS
176	174	95	101	95.5	73749	PASS
177	176	5	9	6.6	4884	PASS