

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010224\
 Data File : VN080625.D
 Acq On : 02 Jan 2024 15:19
 Operator : JC\MD
 Sample : 06038-03 200X
 Misc : 1.10g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T-401

Quant Time: Jan 03 01:46:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122723W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 28 03:07:01 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	287614	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	568509	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	597230	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	291184	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	183585	49.762	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	99.520%		
35) Dibromofluoromethane	8.165	113	167790	45.634	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	91.260%		
50) Toluene-d8	10.571	98	625628	48.733	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	97.460%		
62) 4-Bromofluorobenzene	12.853	95	267180	61.839	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	123.680%#		
Target Compounds						
					Qvalue	
3) Chloromethane	2.365	50	31330	7.579	ug/l	91
40) Benzene	8.606	78	25350	1.896	ug/l	99
67) Ethyl Benzene	11.965	91	186243	8.967	ug/l	98
68) m/p-Xylenes	12.070	106	22380	2.904	ug/l	98
69) o-Xylene	12.400	106	8632	1.140	ug/l	78
73) Isopropylbenzene	12.700	105	21240	1.066	ug/l	96
84) 1,2,4-Trimethylbenzene	13.482	105	62592	3.673	ug/l	98
95) Naphthalene	15.641	128	1039340	68.742	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010224\
Data File : VN080625.D
Acq On : 02 Jan 2024 15:19
Operator : JC\MD
Sample : 06038-03 200X
Misc : 1.10g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T-401

Quant Time: Jan 03 01:46:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122723W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 28 03:07:01 2023
Response via : Initial Calibration

