

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010224\
 Data File : VN080626.D
 Acq On : 02 Jan 2024 15:43
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Jan 03 01:46:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122723W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 28 03:07:01 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	286434	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	560476	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	588614	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	266428	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	184253	50.148	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	100.300%
35) Dibromofluoromethane	8.165	113	168838	46.577	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.160%
50) Toluene-d8	10.565	98	620211	49.004	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.000%
62) 4-Bromofluorobenzene	12.847	95	249563	58.590	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	117.180%
Target Compounds						
						Qvalue
3) Chloromethane	2.365	50	14238	3.006	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010224\
Data File : VN080626.D
Acq On : 02 Jan 2024 15:43
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Jan 03 01:46:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122723W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 28 03:07:01 2023
Response via : Initial Calibration

