



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010225\
Data File : VN085371.D
Acq On : 02 Jan 2025 10:02
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N122024W.M
Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
Last Update : Sat Dec 21 02:03:07 2024



AutoFind: Scans 1851, 1852, 1853; Background Corrected with Scan 1844

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.4	5134	PASS
75	95	30	60	56.7	13019	PASS
95	95	100	100	100.0	22949	PASS
96	95	5	9	7.0	1595	PASS
173	174	0.00	2	2.0	349	PASS
174	95	50	100	76.0	17451	PASS
175	174	5	9	8.2	1431	PASS
176	174	95	101	96.4	16826	PASS
177	176	5	9	6.4	1082	PASS