

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN010319\
 Data File : VN053269.D
 Acq On : 3 Jan 2019 18:18
 Operator : MD\SY
 Sample : PB116111ZHE#14
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB116111ZHE#14

Quant Time: Jan 04 00:56:16 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N121818W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 18 13:03:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1145271	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1783324	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1607436	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	577865	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	568734	48.93	ug/l	0.00
Spiked Amount			Recovery	=	97.86%	
35) Dibromofluoromethane	7.59	113	487988	59.90	ug/l	0.00
Spiked Amount			Recovery	=	119.80%	
50) Toluene-d8	10.09	98	2181435	49.88	ug/l	0.00
Spiked Amount			Recovery	=	99.76%	
62) 4-Bromofluorobenzene	12.40	95	691565	45.05	ug/l	0.00
Spiked Amount			Recovery	=	90.10%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	28823	12.648	ug/l #	87
18) Methyl Acetate	4.33	43	9412	1.884	ug/l	92
20) Methylene Chloride	4.55	84	25567	2.057	ug/l	87
25) 2-Butanone	6.84	43	4379	1.145	ug/l	99
29) Tetrahydrofuran	7.22	42	2434	0.852	ug/l #	66
31) Cyclohexane	7.66	56	18696	Below Cal	#	1
36) 1,1-Dichloropropene	7.67	75	74797	4.370	ug/l #	48
38) Carbon Tetrachloride	7.67	117	99130	6.318	ug/l #	16
43) Isopropyl Acetate	8.17	43	58693	2.175	ug/l #	80
45) 1,2-Dichloropropane	9.27	63	3582	0.270	ug/l #	44
48) Methyl methacrylate	9.18	41	15280	1.648	ug/l #	12
49) 1,4-Dioxane	9.27	88	261	2.225	ug/l #	1
51) 4-Methyl-2-Pentanone	10.02	43	19438	2.101	ug/l #	35
54) cis-1,3-Dichloropropene	9.90	75	131779	6.648	ug/l #	59
57) 1,3-Dichloropropane	10.76	76	26490	1.403	ug/l #	42
59) 2-Hexanone	10.76	43	343690	54.895	ug/l #	50
71) Bromoform	12.40	173	4403	0.492	ug/l #	1
74) N-amyl acetate	11.88	43	39403	3.139	ug/l #	52
76) 1,2,3-Trichloropropane	12.40	75	324865	33.396	ug/l #	100
80) 1,3,5-Trimethylbenzene	12.73	105	18974	0.515	ug/l	98
81) trans-1,4-Dichloro-2-buten	12.40	75	324865	107.638	ug/l #	12
93) 1,2,4-Trichlorobenzene	14.91	180	31	0.599	ug/l #	12
95) Naphthalene	15.14	128	106	1.151	ug/l #	69


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN010319\
Data File : VN053269.D
Acq On : 3 Jan 2019 18:18
Operator : MD\SY
Sample : PB116111ZHE#14
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB116111ZHE#14

Quant Time: Jan 04 00:56:16 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N121818W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 18 13:03:37 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.67 min Scan# 1833

Delta R.T. -0.00 min

Lab File: VN053269.D

Acq: 3 Jan 2019 18:18

Instrument :

MSVOA_N

ClientSampleId :


PB116111ZHE#14

Tgt Ion:168 Resp: 1145271

Ion Ratio Lower Upper

168 100

99 49.2 40.2 60.2



Abundance Ion 167.90 (167.60 to 168.60): VN053269.D (-1740) (-)

Ion 98.90 (98.60 to 99.60): VN053269.D (-1740) (-)

7.67

500000

400000



300000

200000

100000

0

Time-->



#16

Acetone

Concen: 12.648 ug/l

RT: 3.82 min Scan# 637

Delta R.T. 0.01 min

Lab File: VN053269.D


Acq: 3 Jan 2019 18:18

Tgt Ion: 43 Resp: 28823

Ion Ratio Lower Upper

43 100

58 38.9 25.4 38.0#



Abundance Ion 43.00 (42.70 to 43.70): VN053269.D (-541) (-)

Ion 58.00 (57.70 to 58.70): VN053269.D (-541) (-)

3.82

12000

10000

8000


6000


4000

2000

0

Time-->









#18
Methyl Acetate
Concen: 1.884 ug/l
RT: 4.33 min Scan# 795
Delta R.T. 0.01 min
Lab File: VN053269.D
Acq: 3 Jan 2019 18:18




Instrument :
MSVOA_N
ClientSampleId :
PB116111ZHE#14


Tgt Ion: 43 Resp: 9412
Ion Ratio Lower Upper
43 100
74 20.2 19.2 28.8



#20
Methylene Chloride
Concen: 2.057 ug/l
RT: 4.55 min Scan# 865
Delta R.T. -0.00 min
Lab File: VN053269.D
Acq: 3 Jan 2019 18:18

Tgt Ion: 84 Resp: 25567
Ion Ratio Lower Upper
84 100
49 103.7 96.8 145.2
51 33.9 29.9 44.9
86 56.2 51.8 77.6





#25

2-Butanone

Concen: 1.145 ug/l

RT: 6.84 min Scan# 1575

Delta R.T. 0.01 min

Lab File: VN053269.D

Acq: 3 Jan 2019 18:18

Instrument :

MSVOA_N

ClientSampleId :


PB116111ZHE#14

Tgt Ion: 43 Resp: 4379

Ion Ratio Lower Upper




43 100

72 26.7 21.0 31.6



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 72.00 (71.70 to 72.70): VN0



#29

Tetrahydrofuran

Concen: 0.852 ug/l

RT: 7.22 min Scan# 1694

Delta R.T. 0.01 min

Lab File: VN053269.D

Acq: 3 Jan 2019 18:18


Tgt Ion: 42 Resp: 2434

Ion Ratio Lower Upper

42 100

72 30.9 34.7 52.1#



71 9.6 32.3 48.5#




Abundance Ion 42.00 (41.70 to 42.70): VN0

Ion 72.00 (71.70 to 72.70): VN0

Ion 71.00 (70.70 to 71.70): VN0





#33

1,2-Dichloroethane-d4

Concen: Below ug/l

RT: 8.03 min Scan# 1945

Delta R.T. -0.00 min

Lab File: VN053269.D

Acq: 3 Jan 2019 18:18

Instrument :

MSVOA_N

ClientSampleId :


PB116111ZHE#14

Tgt Ion: 65 Resp: 568734

Ion Ratio Lower Upper

65 100

67 55.3 0.0 110.4



Abundance Ion 64.90 (64.60 to 65.60): VN053269.D (-1925) (-)

Ion 66.90 (66.60 to 67.60): VN053269.D (-1925) (-)

8.03

250000

200000



150000

100000

50000

0

Time-->



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.59 min Scan# 2119

Delta R.T. 0.00 min

Lab File: VN053269.D

Acq: 3 Jan 2019 18:18


Tgt Ion: 114 Resp: 1783324

Ion Ratio Lower Upper

114 100

63 19.6 0.0 39.8

88 15.2 0.0 31.0



Abundance Ion 113.90 (113.60 to 114.60): VN053269.D (-2100) (-)

Ion 63.00 (62.70 to 63.70): VN053269.D (-2100) (-)

Ion 87.90 (87.60 to 88.60): VN053269.D (-2100) (-)

8.59

1200000

1000000

800000


600000


400000

200000

0

Time-->





#35

Dibromofluoromethane

Concen: 50.000 ug/l

RT: 7.59 min Scan# 1810

Delta R.T. 0.00 min

Lab File: VN053269.D

Acq: 3 Jan 2019 18:18

Instrument :

MSVOA_N

ClientSampleId :

PB116111ZHE#14


Tgt Ion:113 Resp: 487988

Ion Ratio Lower Upper

113 100

111 101.7 83.0 124.4




192 22.9 17.5 26.3



Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V



#36

1,1-Dichloropropene

Concen: 4.370 ug/l

RT: 7.67 min Scan# 1833

Delta R.T. -0.13 min

Lab File: VN053269.D

Acq: 3 Jan 2019 18:18


Tgt Ion: 75 Resp: 74797

Ion Ratio Lower Upper

75 100

110 8.1 18.2 54.6#



77 0.0 24.9 37.3#




Abundance Ion 74.90 (74.60 to 75.60): V

Ion 109.90 (109.60 to 110.60): V

Ion 76.90 (76.60 to 77.60): V





#38

Carbon Tetrachloride

Concen: 6.318 ug/l

RT: 7.67 min Scan# 1833

Delta R.T. -0.11 min

Lab File: VN053269.D

Acq: 3 Jan 2019 18:18

Instrument :

MSVOA_N

ClientSampleId :

PB116111ZHE#14


Tgt Ion:117 Resp: 99130

Ion Ratio Lower Upper

117 100

119 5.5 77.1 115.7#




121 0.0 24.9 37.3#



Abundance Ion 116.80 (116.50 to 117.50): V

Ion 118.80 (118.50 to 119.50): V

Ion 120.80 (120.50 to 121.50): V



#43

Isopropyl Acetate

Concen: 2.175 ug/l

RT: 8.17 min Scan# 1989

Delta R.T. 0.01 min

Lab File: VN053269.D

Acq: 3 Jan 2019 18:18


Tgt Ion: 43 Resp: 58693

Ion Ratio Lower Upper

43 100

61 15.5 22.9 34.3#



87 9.1 10.6 15.8#




Abundance Ion 43.00 (42.70 to 43.70): V

Ion 61.00 (60.70 to 61.70): V

Ion 87.00 (86.70 to 87.70): V









#45
 1,2-Dichloropropane
 Concen: 0.270 ug/l
 RT: 9.27 min Scan# 2331
 Delta R.T. 0.15 min
 Lab File: VN053269.D
 Acq: 3 Jan 2019 18:18




Instrument :
 MSVOA_N
 ClientSampleId :
 PB116111ZHE#14


Tgt Ion: 63 Resp: 3582
 Ion Ratio Lower Upper
 63 100
 65 0.0 24.6 37.0#



#48
 Methyl methacrylate
 Concen: 1.648 ug/l
 RT: 9.18 min Scan# 2303
 Delta R.T. -0.02 min
 Lab File: VN053269.D
 Acq: 3 Jan 2019 18:18

Tgt Ion: 41 Resp: 15280
 Ion Ratio Lower Upper
 41 100
 69 0.0 72.2 108.2#
 39 0.0 41.9 62.9#





#49

1,4-Dioxane

Concen: 2.225 ug/l

RT: 9.27 min Scan# 2330

Delta R.T. 0.07 min

Lab File: VN053269.D

Acq: 3 Jan 2019 18:18

Instrument :

MSVOA_N

ClientSampleId :

PB116111ZHE#14


Tgt Ion: 88 Resp: 261

Ion Ratio Lower Upper

88 100

43 982262.5 26.6 40.0#




58 4351.0 58.4 87.6#



Abundance Ion 88.00 (87.70 to 88.70): VN0

Ion 43.00 (42.70 to 43.70): VN0

Ion 58.00 (57.70 to 58.70): VN0



#50

Toluene-d8

Concen: 2.225 ug/l

RT: 10.09 min Scan# 2587

Delta R.T. -0.00 min

Lab File: VN053269.D


Acq: 3 Jan 2019 18:18

Tgt Ion: 98 Resp: 2181435

Ion Ratio Lower Upper



98 100


100 64.5 51.6 77.4



Abundance Ion 98.00 (97.70 to 98.70): VN0

Ion 100.00 (99.70 to 100.70): VN0









#51
 4-Methyl-2-Pentanone
 Concen: 2.101 ug/l
 RT: 10.02 min Scan# 2566
 Delta R.T. 0.04 min
 Lab File: VN053269.D
 Acq: 3 Jan 2019 18:18




Instrument :
 MSVOA_N
 ClientSampleId :
 PB116111ZHE#14


Tgt Ion: 43 Resp: 19438
 Ion Ratio Lower Upper
 43 100
 58 0.0 32.2 48.4#



#54
 cis-1,3-Dichloropropene
 Concen: 6.648 ug/l
 RT: 9.90 min Scan# 2529
 Delta R.T. 0.07 min
 Lab File: VN053269.D
 Acq: 3 Jan 2019 18:18

Tgt Ion: 75 Resp: 131779
 Ion Ratio Lower Upper
 75 100
 77 0.5 25.8 38.8#
 39 25.2 34.8 52.2#









#57
 1,3-Dichloropropane
 Concen: 1.403 ug/l
 RT: 10.76 min Scan# 2796
 Delta R.T. 0.06 min
 Lab File: VN053269.D
 Acq: 3 Jan 2019 18:18




Instrument :
 MSVOA_N
 ClientSampleId :
 PB116111ZHE#14


Tgt Ion: 76 Resp: 26490
 Ion Ratio Lower Upper
 76 100
 78 0.0 25.8 38.8#



#59
 2-Hexanone
 Concen: 54.895 ug/l
 RT: 10.76 min Scan# 2796
 Delta R.T. 0.01 min
 Lab File: VN053269.D
 Acq: 3 Jan 2019 18:18

Tgt Ion: 43 Resp: 343690
 Ion Ratio Lower Upper
 43 100
 58 20.0 28.1 84.3#





#62

4-Bromofluorobenzene

Concen: N.D. ug/l

RT: 12.40 min Scan# 3305

Delta R.T. -0.00 min

Lab File: VN053269.D

Acq: 3 Jan 2019 18:18

Instrument :

MSVOA_N

ClientSampleId :

PB116111ZHE#14


Tgt Ion: 95 Resp: 691565

Ion Ratio Lower Upper

95 100

174 90.1 0.0 178.8

176 87.4 0.0 173.8



Abundance Ion 94.90 (94.60 to 95.60): VN052954.D (-3290) (-)

Ion 173.80 (173.50 to 174.50): VN052954.D (-3290) (-)

Ion 175.80 (175.50 to 176.50): VN052954.D (-3290) (-)

12.40

500000

400000



300000

200000

100000

0

Time-->



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.41 min Scan# 2996

Delta R.T. -0.00 min

Lab File: VN053269.D

Acq: 3 Jan 2019 18:18


Tgt Ion: 117 Resp: 1607436

Ion Ratio Lower Upper

117 100

82 53.5 42.8 64.2

119 32.0 25.5 38.3



Abundance Ion 116.90 (116.60 to 117.60): VN052954.D (-2980) (-)

Ion 82.00 (81.70 to 82.70): VN052954.D (-2980) (-)

Ion 118.90 (118.60 to 119.60): VN052954.D (-2980) (-)

11.41

1200000

1000000

800000


600000


400000

200000

0

Time-->





#71

Bromoform

Concen: 0.492 ug/l

RT: 12.40 min Scan# 3305

Delta R.T. 0.27 min

Lab File: VN053269.D

Acq: 3 Jan 2019 18:18

Instrument :

MSVOA_N

ClientSampleId :

PB116111ZHE#14


Tgt Ion:173 Resp: 4403

Ion Ratio Lower Upper

173 100

175 921.1 24.3 72.9#




254 0.0 0.1 0.1#



Abundance Ion 172.70 (172.40 to 173.40): V

Ion 174.70 (174.40 to 175.40): V

Ion 254.40 (254.10 to 255.10): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.35 min Scan# 3599

Delta R.T. 0.00 min

Lab File: VN053269.D

Acq: 3 Jan 2019 18:18


Tgt Ion:152 Resp: 577865

Ion Ratio Lower Upper

152 100

115 56.4 28.3 85.0



150 157.5 0.0 347.8




Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V





#74

N-ethyl acetate

Concen: 3.139 ug/l

RT: 11.88 min Scan# 3142

Delta R.T. -0.19 min

Lab File: VN053269.D

Acq: 3 Jan 2019 18:18

Instrument :

MSVOA_N

ClientSampleId :

PB116111ZHE#14

Tgt Ion: 43 Resp: 39403


Ion Ratio Lower Upper

43 100

70 2.3 34.2 51.4#

55 4.4 20.3 30.5#

61 11.8 20.0 30.0#






Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 70.00 (69.70 to 70.70): VN0

Ion 55.00 (54.70 to 55.70): VN0

Ion 61.00 (60.70 to 61.70): VN0



#76

1,2,3-Trichloropropane

Concen: 33.396 ug/l

RT: 12.40 min Scan# 3305

Delta R.T. -0.15 min

Lab File: VN053269.D


Acq: 3 Jan 2019 18:18

Tgt Ion: 75 Resp: 324865

Ion Ratio Lower Upper



75 100


77 1.3 0.0 0.0#



Abundance Ion 74.90 (74.60 to 75.60): VN0

Ion 77.00 (76.70 to 77.70): VN0









#80
 1,3,5-Trimethylbenzene
 Concen: 0.515 ug/l
 RT: 12.73 min Scan# 3409
 Delta R.T. 0.00 min
 Lab File: VN053269.D
 Acq: 3 Jan 2019 18:18




Instrument :
 MSVOA_N
 ClientSampleId :
 PB116111ZHE#14


Tgt Ion: 105 Resp: 18974
 Ion Ratio Lower Upper
 105 100
 120 47.3 24.4 73.2



#81
 trans-1,4-Dichloro-2-butene
 Concen: 107.638 ug/l
 RT: 12.40 min Scan# 3305
 Delta R.T. 0.10 min
 Lab File: VN053269.D
 Acq: 3 Jan 2019 18:18

Tgt Ion: 75 Resp: 324865
 Ion Ratio Lower Upper
 75 100
 53 0.0 74.7 112.1#
 89 0.0 36.1 54.1#







#93
 1,2,4-Trichlorobenzene
 Concen: 0.599 ug/l
 RT: 14.91 min Scan# 4086
 Delta R.T. 0.00 min
 Lab File: VN053269.D
 Acq: 3 Jan 2019 18:18

Instrument :
 MSVOA_N
 ClientSampleId :
 PB116111ZHE#14



Tgt Ion:180 Resp: 31
 Ion Ratio Lower Upper
 180 100
 182 0.0 47.8 143.3#
 145 0.0 14.2 42.6#



Abundance Ion 179.80 (179.50 to 180.50): V
 Ion 181.80 (181.50 to 182.50): V
 Ion 144.80 (144.50 to 145.50): V




Time--> 14.90 14.91 14.92




#95
 Naphthalene
 Concen: 1.151 ug/l
 RT: 15.14 min Scan# 4156
 Delta R.T. 0.00 min
 Lab File: VN053269.D
 Acq: 3 Jan 2019 18:18

Tgt Ion:128 Resp: 106
 Ion Ratio Lower Upper
 128 100
 127 0.0 10.2 15.4#
 129 0.0 8.6 13.0#



Abundance Ion 128.00 (127.70 to 128.70): V
 Ion 127.00 (126.70 to 127.70): V
 Ion 129.00 (128.70 to 129.70): V



Time--> 15.12 15.13 15.14 15.15

