

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN010319\
 Data File : VN053265.D
 Acq On : 3 Jan 2019 16:31
 Operator : MD\SY
 Sample : PB116111ZHE#10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB116111ZHE#10

Quant Time: Jan 04 07:36:25 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N121818W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 18 13:03:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1148963	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1785422	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1656717	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	624929	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	576406	49.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.88%	
35) Dibromofluoromethane	7.59	113	490073	60.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	120.18%	
50) Toluene-d8	10.09	98	2214462	50.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.16%	
62) 4-Bromofluorobenzene	12.40	95	730752	47.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.08%	
Target Compounds						
16) Acetone	3.82	43	33817	14.792	ug/l	99
20) Methylene Chloride	4.55	84	24181	1.939	ug/l	94
43) Isopropyl Acetate	8.17	43	52590	1.834	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN010319\
Data File : VN053265.D
Acq On : 3 Jan 2019 16:31
Operator : MD\SY
Sample : PB116111ZHE#10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB116111ZHE#10

Quant Time: Jan 04 07:36:25 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N121818W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 18 13:03:37 2018
Response via : Initial Calibration

