

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN010319\
 Data File : VN053272.D
 Acq On : 3 Jan 2019 19:38
 Operator : MD\SY
 Sample : K1002-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 25 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 AU-05-010219-A

Quant Time: Jan 04 07:58:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N121818W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 18 13:03:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1033113	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1598908	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1438093	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	532163	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	523379	49.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.84%	
35) Dibromofluoromethane	7.59	113	448569	61.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	122.84%	
50) Toluene-d8	10.09	98	1966836	50.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.32%	
62) 4-Bromofluorobenzene	12.40	95	627506	45.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.18%	
Target Compounds						
16) Acetone	3.81	43	18574	9.036	ug/l	97
20) Methylene Chloride	4.55	84	515635	45.985	ug/l	98
25) 2-Butanone	6.83	43	12066	3.498	ug/l	94
43) Isopropyl Acetate	8.17	43	47037	1.831	ug/l #	83
95) Naphthalene	15.13	128	5235	1.391	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN010319\
Data File : VN053272.D
Acq On : 3 Jan 2019 19:38
Operator : MD\SY
Sample : K1002-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 25 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
AU-05-010219-A

Quant Time: Jan 04 07:58:27 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N121818W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 18 13:03:37 2018
Response via : Initial Calibration

