

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010322\
 Data File : VN070436.D
 Acq On : 3 Jan 2022 14:45
 Operator : JC/MD
 Sample : M5177-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6RE

Quant Time: Jan 04 01:33:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 01/04/2022
 Supervised By : Mahesh Dadoda 01/04/2022

01/04/2022
 Supervised By : Mahesh
 Dadoda

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.083	168	942043	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	1609488	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1396958	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	456245	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	681496	50.314	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	100.620%
35) Dibromofluoromethane	8.021	113	476494	48.930	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.860%
50) Toluene-d8	10.440	98	1890699	47.527	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.060%
62) 4-Bromofluorobenzene	12.731	95	646326	44.983	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	89.960%
Target Compounds						
16) Acetone	4.293	43	21331	2.426	ug/l	98
20) Methylene Chloride	5.093	84	6270m	0.515	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010322\
Data File : VN070436.D
Acq On : 3 Jan 2022 14:45
Operator : JC/MD
Sample : M5177-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6RE

Quant Time: Jan 04 01:33:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 01/04/2022
Supervised By : Mahesh Dadoda 01/04/2022

01/04/2022
Supervised By : Mahesh
Dadoda

01/04/2022

