



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010322\
Data File : VN070426.D
Acq On : 3 Jan 2022 9:52
Operator : JC/MD
Sample : BFB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Title : SW846 8260
Last Update : Mon Dec 27 18:47:12 2021



AutoFind: Scans 4004, 4005, 4006; Background Corrected with Scan 3985

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.0	57467	PASS
75	95	30	60	51.9	142187	PASS
95	95	100	100	100.0	273941	PASS
96	95	5	9	6.9	18877	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	77.1	211221	PASS
175	174	5	9	7.1	15002	PASS
176	174	95	101	97.4	205632	PASS
177	176	5	9	6.2	12683	PASS