

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010323\
 Data File : VN076063.D
 Acq On : 03 Jan 2023 15:02
 Operator : JC\MD
 Sample : N6232-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-11

Quant Time: Jan 04 04:18:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	352750	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	551241	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	486322	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	199821	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	179485	46.770	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.540%
35) Dibromofluoromethane	8.177	113	163765	47.972	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.940%
50) Toluene-d8	10.571	98	635398	49.782	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.560%
62) 4-Bromofluorobenzene	12.853	95	196251	46.423	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.840%
Target Compounds						
					Qvalue	
19) Methyl tert-butyl Ether	5.812	73	16009	1.555	ug/l	100
84) 1,2,4-Trimethylbenzene	13.488	105	4332	0.361	ug/l	98
95) Naphthalene	15.647	128	8650	2.938	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010323\
Data File : VN076063.D
Acq On : 03 Jan 2023 15:02
Operator : JC\MD
Sample : N6232-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-11

Quant Time: Jan 04 04:18:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

