

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010323\
 Data File : VN076065.D
 Acq On : 03 Jan 2023 15:50
 Operator : JC\MD
 Sample : N6232-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-13

Quant Time: Jan 04 04:19:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	328647	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	514948	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	441189	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	182463	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	166160	46.473	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.940%
35) Dibromofluoromethane	8.177	113	154427	48.425	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.840%
50) Toluene-d8	10.571	98	584358	49.010	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.020%
62) 4-Bromofluorobenzene	12.853	95	183827	46.549	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.100%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	5.542	59	36976	73.088	ug/l #	72
19) Methyl tert-butyl Ether	5.806	73	24145737	2517.292	ug/l	100
22) Diisopropyl ether	6.689	45	14735	1.781	ug/l	94
67) Ethyl Benzene	11.965	91	12750	0.834	ug/l	95
68) m/p-Xylenes	12.076	106	6027	0.996	ug/l	90
69) o-Xylene	12.406	106	1317	0.220	ug/l	93
84) 1,2,4-Trimethylbenzene	13.488	105	6667	0.609	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010323\
 Data File : VN076065.D
 Acq On : 03 Jan 2023 15:50
 Operator : JC\MD
 Sample : N6232-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-13

Quant Time: Jan 04 04:19:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

