

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010323\
 Data File : VN076069.D
 Acq On : 03 Jan 2023 17:27
 Operator : JC\MD
 Sample : N6232-07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-1

Quant Time: Jan 04 04:20:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	360999	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	557801	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	479749	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	198981	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	184471	46.971	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.940%
35) Dibromofluoromethane	8.177	113	165996	48.054	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.100%
50) Toluene-d8	10.571	98	636275	49.264	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.520%
62) 4-Bromofluorobenzene	12.853	95	198586	46.423	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.840%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010323\
Data File : VN076069.D
Acq On : 03 Jan 2023 17:27
Operator : JC\MD
Sample : N6232-07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-1

Quant Time: Jan 04 04:20:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

