

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010324\
 Data File : VN080633.D
 Acq On : 03 Jan 2024 12:22
 Operator : JC\MD
 Sample : VN0103WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0103WBL01

Quant Time: Jan 04 01:56:25 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122723W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 28 03:07:01 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	301739	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	577993	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	600219	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	219383	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	177742	45.923	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	91.840%
35) Dibromofluoromethane	8.165	113	173759	46.482	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.960%
50) Toluene-d8	10.571	98	616038	47.199	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.400%
62) 4-Bromofluorobenzene	12.847	95	202292	46.053	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.100%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010324\
Data File : VN080633.D
Acq On : 03 Jan 2024 12:22
Operator : JC\MD
Sample : VN0103WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0103WBL01

Quant Time: Jan 04 01:56:25 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122723W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 28 03:07:01 2023
Response via : Initial Calibration

