

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010324\
 Data File : VN080639.D
 Acq On : 03 Jan 2024 14:56
 Operator : JC\MD
 Sample : PB158184ZHE#02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB158184ZHE#02

Quant Time: Jan 04 01:59:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122723W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 28 03:07:01 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	308647	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	611826	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	667164	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	290482	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	197486	49.882	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.760%
35) Dibromofluoromethane	8.165	113	177288	44.803	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.600%
50) Toluene-d8	10.565	98	678965	49.143	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.280%
62) 4-Bromofluorobenzene	12.847	95	269271	57.911	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	115.820%
Target Compounds						
					Qvalue	
16) Acetone	4.424	43	36966	34.348	ug/l	89
20) Methylene Chloride	5.277	84	25014	7.284	ug/l	93
25) 2-Butanone	7.483	43	18383	9.139	ug/l #	81
37) Ethyl Acetate	7.553	43	4597	1.000	ug/l #	88
43) Isopropyl Acetate	8.688	43	164475	22.082	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010324\
Data File : VN080639.D
Acq On : 03 Jan 2024 14:56
Operator : JC\MD
Sample : PB158184ZHE#02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB158184ZHE#02

Quant Time: Jan 04 01:59:58 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122723W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 28 03:07:01 2023
Response via : Initial Calibration

