

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN010419\
 Data File : VN053282.D
 Acq On : 4 Jan 2019 15:17
 Operator : MD\SY
 Sample : K1046-02 5X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 002M-35TH-AVE(JAN)

Quant Time: Jan 05 00:19:35 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N122818W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Dec 28 10:33:50 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.19	128	135865	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.58	114	773567	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	690416	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	8.02	65	260640	30.20	ug/l	0.00
Spiked Amount 30.000	Range 50 - 169		Recovery	=	100.67%	
60) 4-Bromofluorobenzene	12.40	95	312820	29.61	ug/l	0.00
Spiked Amount 30.000	Range 56 - 143		Recovery	=	98.70%	
63) Toluene-d8	10.09	98	906433	29.14	ug/l	0.00
Spiked Amount 30.000	Range 66 - 137		Recovery	=	97.13%	

Target Compounds

					Ovalue
15) Acetone	3.82	58	133212	227.069 ug/l	97
58) 4-Methyl-2-Pentanone	9.99	43	3010	0.458 ug/l	98
62) Toluene	10.16	91	92144	2.434 ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN010419\
Data File : VN053282.D
Acq On : 4 Jan 2019 15:17
Operator : MD\SY
Sample : K1046-02 5X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
002M-35TH-AVE(JAN)

Quant Time: Jan 05 00:19:35 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N122818W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Dec 28 10:33:50 2018
Response via : Initial Calibration

