


Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN010419\
 Data File : VN053275.D
 Acq On : 4 Jan 2019 8:24
 Operator : MD\SY
 Sample : BFB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 1 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT3.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\624N122818W.M
 Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 Last Update : Fri Dec 28 10:33:50 2018



AutoFind: Scans 3304, 3305, 3306; Background Corrected with Scan 3291

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.1	13574	PASS
75	95	30	60	44.3	37309	PASS
95	95	100	100	100.0	84208	PASS
96	95	5	9	6.3	5269	PASS
173	174	0.00	2	0.9	677	PASS
174	95	50	100	89.4	75280	PASS
175	174	5	9	6.7	5063	PASS
176	174	95	101	97.2	73178	PASS
177	176	5	9	6.9	5014	PASS