

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010421\
 Data File : VN065376.D
 Acq On : 4 Jan 2021 15:08
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 05 00:58:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220w.M
 Quant Title : Sw846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	123	0.00
2 T	Dichlorodifluoromethane	50.000	46.602	6.8	108	0.00
3 P	Chloromethane	50.000	51.116	-2.2	121	0.00
4 C	Vinyl Chloride	50.000	52.922	-5.8#	124	0.00
5 T	Bromomethane	50.000	52.634	-5.3	127	0.00
6 T	Chloroethane	50.000	52.514	-5.0	123	0.00
7 T	Trichlorofluoromethane	50.000	46.995	6.0	112	0.00
8 T	Diethyl Ether	50.000	49.108	1.8	121	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	48.846	2.3	119	0.00
10 T	Methyl Iodide	50.000	49.599	0.8	115	0.00
11 T	Tert butyl alcohol	250.000	224.546	10.2	107	0.00
12 CM	1,1-Dichloroethene	50.000	47.104	5.8#	117	0.00
13 T	Acrolein	250.000	219.746	12.1	106	0.00
14 T	Allyl chloride	50.000	45.091	9.8	108	0.00
15 T	Acrylonitrile	250.000	248.241	0.7	117	0.00
16 T	Acetone	250.000	190.657	23.7	92	0.00
17 T	Carbon Disulfide	50.000	46.120	7.8	108	0.00
18 T	Methyl Acetate	50.000	49.796	0.4	119	0.00
19 T	Methyl tert-butyl Ether	50.000	47.204	5.6	114	0.00
20 T	Methylene Chloride	50.000	47.208	5.6	123	0.00
21 T	trans-1,2-Dichloroethene	50.000	49.755	0.5	119	0.00
22 T	Diisopropyl ether	50.000	47.929	4.1	117	0.00
23 T	Vinyl Acetate	250.000	232.301	7.1	110	0.00
24 P	1,1-Dichloroethane	50.000	50.242	-0.5	120	0.00
25 T	2-Butanone	250.000	222.333	11.1	105	0.00
26 T	2,2-Dichloropropane	50.000	44.193	11.6	109	0.00
27 T	cis-1,2-Dichloroethene	50.000	50.930	-1.9	122	0.00
28 T	Bromochloromethane	50.000	45.245	9.5	118	0.00
29 T	Tetrahydrofuran	250.000	223.138	10.7	108	0.00
30 C	Chloroform	50.000	49.450	1.1#	119	0.00
31 T	Cyclohexane	50.000	44.384	11.2	112	0.00
32 T	1,1,1-Trichloroethane	50.000	46.917	6.2	113	0.00
33 S	1,2-Dichloroethane-d4	50.000	46.325	7.3	112	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	120	0.00
35 S	Dibromofluoromethane	50.000	51.254	-2.5	122	0.00
36 T	1,1-Dichloropropene	50.000	50.603	-1.2	120	0.00
37 T	Ethyl Acetate	50.000	45.788	8.4	107	0.00
38 T	Carbon Tetrachloride	50.000	47.361	5.3	112	0.00
39 T	Methylcyclohexane	50.000	48.894	2.2	113	0.00
40 TM	Benzene	50.000	51.696	-3.4	123	0.00
41 T	Methacrylonitrile	50.000	43.386	13.2	108	0.00
42 TM	1,2-Dichloroethane	50.000	46.743	6.5	112	0.00
43 T	Isopropyl Acetate	50.000	44.795	10.4	104	0.00
44 TM	Trichloroethene	50.000	51.780	-3.6	123	0.00
45 C	1,2-Dichloropropane	50.000	52.479	-5.0#	123	0.00
46 T	Dibromomethane	50.000	50.931	-1.9	121	0.00
47 T	Bromodichloromethane	50.000	49.380	1.2	117	0.00
48 T	Methyl methacrylate	50.000	45.703	8.6	108	0.00
49 T	1,4-Dioxane	1000.000	1073.384	-7.3	127	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010421\
 Data File : VN065376.D
 Acq On : 4 Jan 2021 15:08
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 05 00:58:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220w.M
 Quant Title : Sw846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)

50 S	Toluene-d8	50.000	52.056	-4.1	124	0.00
51 T	4-Methyl-2-Pentanone	250.000	235.000	6.0	110	0.00
52 CM	Toluene	50.000	52.858	-5.7#	125	0.00
53 T	t-1,3-Dichloropropene	50.000	49.441	1.1	116	0.00
54 T	cis-1,3-Dichloropropene	50.000	50.069	-0.1	120	0.00
55 T	1,1,2-Trichloroethane	50.000	52.728	-5.5	124	0.00
56 T	Ethyl methacrylate	50.000	50.262	-0.5	118	0.00
57 T	1,3-Dichloropropane	50.000	53.204	-6.4	124	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	233.132	6.7	112	0.00
59 T	2-Hexanone	250.000	233.221	6.7	107	0.00
60 T	Dibromochloromethane	50.000	51.015	-2.0	120	0.00
61 T	1,2-Dibromoethane	50.000	51.729	-3.5	122	0.00
62 S	4-Bromofluorobenzene	50.000	49.860	0.3	126	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	124	0.00
64 T	Tetrachloroethene	50.000	49.463	1.1	125	0.00
65 PM	Chlorobenzene	50.000	52.525	-5.0	128	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	51.330	-2.7	125	0.00
67 C	Ethyl Benzene	50.000	50.966	-1.9#	125	0.00
68 T	m/p-Xylenes	100.000	104.818	-4.8	128	0.00
69 T	o-Xylene	50.000	51.935	-3.9	127	0.00
70 T	Styrene	50.000	53.524	-7.0	129	0.00
71 P	Bromoform	50.000	50.171	-0.3	120	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	132	0.00
73 T	Isopropylbenzene	50.000	46.482	7.0	126	0.00
74 T	N-aryl acetate	50.000	42.561	14.9	110	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	48.297	3.4	130	0.00
76 T	1,2,3-Trichloropropane	50.000	49.351	1.3	139	0.00
77 T	Bromobenzene	50.000	49.406	1.2	133	0.00
78 T	n-propylbenzene	50.000	46.771	6.5	125	0.00
79 T	2-Chlorotoluene	50.000	47.399	5.2	126	0.00
80 T	1,3,5-Trimethylbenzene	50.000	47.064	5.9	127	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	43.681	12.6	114	0.00
82 T	4-Chlorotoluene	50.000	48.136	3.7	129	0.00
83 T	tert-Butylbenzene	50.000	47.375	5.3	128	0.00
84 T	1,2,4-Trimethylbenzene	50.000	48.519	3.0	130	0.00
85 T	sec-Butylbenzene	50.000	47.065	5.9	127	0.00
86 T	p-Isopropyltoluene	50.000	47.458	5.1	127	0.00
87 T	1,3-Dichlorobenzene	50.000	51.350	-2.7	135	0.00
88 T	1,4-Dichlorobenzene	50.000	50.883	-1.8	134	0.00
89 T	n-Butylbenzene	50.000	46.602	6.8	122	0.00
90 T	Hexachloroethane	50.000	48.009	4.0	124	0.00
91 T	1,2-Dichlorobenzene	50.000	52.629	-5.3	137	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	45.019	10.0	114	0.00
93 T	1,2,4-Trichlorobenzene	50.000	48.806	2.4	134	0.00
94 T	Hexachlorobutadiene	50.000	46.166	7.7	129	0.00
95 T	Naphthalene	50.000	45.520	9.0	128	0.00
96 T	1,2,3-Trichlorobenzene	50.000	47.907	4.2	135	0.02

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010421\
Data File : VN065376.D
Acq On : 4 Jan 2021 15:08
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jan 05 00:58:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220w.M
Quant Title : Sw846 8260
QLast Update : Tue Dec 22 15:42:18 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 6