

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010422\
 Data File : VN070461.D
 Acq On : 4 Jan 2022 14:11
 Operator : JC/MD
 Sample : M5338-10DL 20X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TW-3DL

Quant Time: Jan 05 03:16:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.083	168	1010261	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	1687570	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1466540	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	496538	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	719071	49.504	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	=	99.000%	
35) Dibromofluoromethane	8.021	113	514821	50.420	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	=	100.840%	
50) Toluene-d8	10.440	98	1963394	47.071	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	=	94.140%	
62) 4-Bromofluorobenzene	12.731	95	705022	46.798	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	=	93.600%	
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	4.186	96	68104	6.619	ug/l	96
16) Acetone	4.296	43	6809	0.722	ug/l	92
20) Methylene Chloride	5.095	84	7878	0.604	ug/l	89
24) 1,1-Dichloroethane	6.385	63	146075	6.200	ug/l	98
27) cis-1,2-Dichloroethene	7.324	96	18497	1.423	ug/l	95
30) Chloroform	7.817	83	583766	25.684	ug/l	99
32) 1,1,1-Trichloroethane	8.013	97	1724009	89.961	ug/l	96
38) Carbon Tetrachloride	8.204	117	213607	13.244	ug/l	98
44) Trichloroethene	9.215	130	7358	0.598	ug/l	93
64) Tetrachloroethene	10.977	164	34301	3.230	ug/l	92
67) Ethyl Benzene	11.843	91	21475	0.377	ug/l	99
68) m/p-Xylenes	11.953	106	27137	1.308	ug/l	95
69) o-Xylene	12.275	106	6810	0.330	ug/l	88
84) 1,2,4-Trimethylbenzene	13.366	105	27924	0.785	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010422\
Data File : VN070461.D
Acq On : 4 Jan 2022 14:11
Operator : JC/MD
Sample : M5338-10DL 20X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TW-3DL

Quant Time: Jan 05 03:16:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

