

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010422\
 Data File : VN070464.D
 Acq On : 4 Jan 2022 15:27
 Operator : JC/MD
 Sample : N1003-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Jan 05 03:17:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	954848	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1614632	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1394757	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	463526	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	694440	50.582	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	101.160%
35) Dibromofluoromethane	8.021	113	478504	48.980	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.960%
50) Toluene-d8	10.440	98	1914440	47.970	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.940%
62) 4-Bromofluorobenzene	12.731	95	671002	46.551	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	93.100%
Target Compounds						
						Qvalue
16) Acetone	4.291	43	16246	1.823	ug/l	98
20) Methylene Chloride	5.101	84	6921	0.561	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010422\
Data File : VN070464.D
Acq On : 4 Jan 2022 15:27
Operator : JC/MD
Sample : N1003-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Jan 05 03:17:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

