

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010422\
 Data File : VN070466.D
 Acq On : 4 Jan 2022 16:17
 Operator : JC/MD
 Sample : N1003-04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANAG

Quant Time: Jan 05 03:17:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	958561	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	1640295	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1417279	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	473149	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	797623	57.873	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 115.740%	
35) Dibromofluoromethane	8.021	113	554172	55.838	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 111.680%	
50) Toluene-d8	10.441	98	2200436	54.274	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 108.540%	
62) 4-Bromofluorobenzene	12.731	95	764108	52.181	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 104.360%	
Target Compounds						
						Qvalue
16) Acetone	4.291	43	19182	2.144	ug/l	94
20) Methylene Chloride	5.093	84	4123	0.333	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010422\
Data File : VN070466.D
Acq On : 4 Jan 2022 16:17
Operator : JC/MD
Sample : N1003-04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANAG

Quant Time: Jan 05 03:17:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

