

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010422\
 Data File : VN070474.D
 Acq On : 4 Jan 2022 19:40
 Operator : JC/MD
 Sample : N1012-03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 31235

Quant Time: Jan 05 03:21:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	957984	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	1623563	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1449098	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.672	152	530261	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	784657	56.967	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 113.940%	
35) Dibromofluoromethane	8.019	113	539273	54.897	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 109.800%	
50) Toluene-d8	10.440	98	2208156	55.026	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 110.060%	
62) 4-Bromofluorobenzene	12.728	95	837160	57.759	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 115.520%	
Target Compounds						
					Qvalue	
16) Acetone	4.288	43	240515	26.899	ug/l	97
20) Methylene Chloride	5.095	84	11287	0.912	ug/l #	85
25) 2-Butanone	7.340	43	90444	8.033	ug/l #	87
39) Methylcyclohexane	9.461	83	8317	0.423	ug/l #	93
40) Benzene	8.464	78	32123	0.645	ug/l	99
95) Naphthalene	15.504	128	38138	3.319	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010422\
Data File : VN070474.D
Acq On : 4 Jan 2022 19:40
Operator : JC/MD
Sample : N1012-03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
31235

Quant Time: Jan 05 03:21:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

