

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010422\
 Data File : VN070476.D
 Acq On : 4 Jan 2022 20:30
 Operator : JC/MD
 Sample : N1012-05
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 31237

Quant Time: Jan 05 03:22:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1043324	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	1755623	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1558453	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.672	152	548614	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	819127	54.605	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 109.200%	
35) Dibromofluoromethane	8.021	113	563387	53.038	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 106.080%	
50) Toluene-d8	10.438	98	2294271	52.871	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 105.740%	
62) 4-Bromofluorobenzene	12.728	95	834637	53.254	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 106.500%	
Target Compounds						
					Qvalue	
16) Acetone	4.291	43	241142	24.763	ug/l	100
20) Methylene Chloride	5.098	84	9418	0.699	ug/l #	83
25) 2-Butanone	7.343	43	23538	1.920	ug/l	97
39) Methylcyclohexane	9.456	83	8117	0.382	ug/l #	80
40) Benzene	8.456	78	43725	0.812	ug/l	98
52) Toluene	10.505	92	14029	0.430	ug/l	96
95) Naphthalene	15.507	128	8363	2.119	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010422\
Data File : VN070476.D
Acq On : 4 Jan 2022 20:30
Operator : JC/MD
Sample : N1012-05
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
31237

Quant Time: Jan 05 03:22:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

