

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010423\
 Data File : VN076079.D
 Acq On : 04 Jan 2023 14:17
 Operator : JC\MD
 Sample : PB150013TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB150013TB

Quant Time: Jan 05 01:23:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	341518	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	546758	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	484400	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	193951	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	181363	48.814	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.620%
35) Dibromofluoromethane	8.177	113	163726	48.354	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.700%
50) Toluene-d8	10.571	98	640303	50.577	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.160%
62) 4-Bromofluorobenzene	12.853	95	195493	46.623	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.240%
Target Compounds						
					Qvalue	
16) Acetone	4.448	43	41976	32.774	ug/l	92
37) Ethyl Acetate	7.577	43	33381	9.239	ug/l	99
43) Isopropyl Acetate	8.694	43	168735	28.312	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010423\
Data File : VN076079.D
Acq On : 04 Jan 2023 14:17
Operator : JC\MD
Sample : PB150013TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB150013TB

Quant Time: Jan 05 01:23:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

