

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010423\
 Data File : VN076080.D
 Acq On : 04 Jan 2023 14:41
 Operator : JC\MD
 Sample : PB150013ZHE#01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB150013ZHE#01

Quant Time: Jan 05 01:23:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	342574	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	549882	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	498453	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	209321	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	184006	49.373	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.740%
35) Dibromofluoromethane	8.171	113	164242	48.231	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.460%
50) Toluene-d8	10.571	98	651741	51.188	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.380%
62) 4-Bromofluorobenzene	12.853	95	209273	49.626	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.260%
Target Compounds						
					Qvalue	
16) Acetone	4.447	43	41610	32.388	ug/l	99
20) Methylene Chloride	5.294	84	9999	2.298	ug/l	95
25) 2-Butanone	7.494	43	10122	5.627	ug/l	90
37) Ethyl Acetate	7.571	43	30733	8.458	ug/l #	90
43) Isopropyl Acetate	8.694	43	186416	31.101	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010423\
Data File : VN076080.D
Acq On : 04 Jan 2023 14:41
Operator : JC\MD
Sample : PB150013ZHE#01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB150013ZHE#01

Quant Time: Jan 05 01:23:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

