

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010423\
 Data File : VN076087.D
 Acq On : 04 Jan 2023 17:30
 Operator : JC\MD
 Sample : PB150013ZHE#08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB150013ZHE#08

Manual Integrations
 APPROVED

Reviewed By : John Carlone 01/05/2023
 Supervised By : Mahesh Dadoda 01/05/2023

Quant Time: Jan 05 01:26:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	306233	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	513172	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	452471	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	179426	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.589	65	166044	49.840	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	99.680%		
35) Dibromofluoromethane	8.177	113	151152	47.562	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	95.120%		
50) Toluene-d8	10.571	98	593276	49.930	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	99.860%		
62) 4-Bromofluorobenzene	12.853	95	180500	45.864	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	91.720%		
Target Compounds						
16) Acetone	4.448	43	41498	36.134	ug/l	98
18) Methyl Acetate	5.059	43	3786	1.115	ug/l #	85
20) Methylene Chloride	5.289	84	7598	1.857	ug/l #	86
25) 2-Butanone	7.494	43	7598	4.725	ug/l	96
37) Ethyl Acetate	7.571	43	32879	9.696	ug/l	98
43) Isopropyl Acetate	8.694	43	269471m	48.174	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010423\
Data File : VN076087.D
Acq On : 04 Jan 2023 17:30
Operator : JC\MD
Sample : PB150013ZHE#08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB150013ZHE#08

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/05/2023
Supervised By :Mahesh Dadoda 01/05/2023

Quant Time: Jan 05 01:26:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

