

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010423\
 Data File : VN076088.D
 Acq On : 04 Jan 2023 17:54
 Operator : JC\MD
 Sample : N6239-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-1S

Quant Time: Jan 05 01:26:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 01/05/2023
 Supervised By : Mahesh Dadoda 01/05/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	322732	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	527115	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	469487	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	187623	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	172407	49.104	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	98.200%		
35) Dibromofluoromethane	8.177	113	155965	47.778	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	95.560%		
50) Toluene-d8	10.571	98	612255	50.164	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	100.320%		
62) 4-Bromofluorobenzene	12.853	95	189022	46.759	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	93.520%		
Target Compounds						
16) Acetone	4.453	43	37258	30.784	ug/l	99
20) Methylene Chloride	5.289	84	9527	2.332	ug/l	95
37) Ethyl Acetate	7.577	43	31587	9.069	ug/l	100
43) Isopropyl Acetate	8.694	43	287816m	50.092	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010423\
Data File : VN076088.D
Acq On : 04 Jan 2023 17:54
Operator : JC\MD
Sample : N6239-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-1S

Manual Integrations
APPROVED

Reviewed By : John Carlone 01/05/2023
Supervised By : Mahesh Dadoda 01/05/2023

Quant Time: Jan 05 01:26:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

