

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010521\
 Data File : VN065379.D
 Acq On : 5 Jan 2021 11:11
 Operator : JC/MD
 Sample : VN0105MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0105MBL01

Quant Time: Jan 06 01:14:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.627	168	274495	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	461154	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.380	117	473710	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.318	152	182138	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.991	65	152307	49.15	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.30%
35) Dibromofluoromethane	7.550	113	140345	50.44	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.88%
50) Toluene-d8	10.058	98	568473	54.52	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	109.04%
62) 4-Bromofluorobenzene	12.376	95	208114	50.42	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	100.84%

Target Compounds	Qvalue
-----	-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010521\
Data File : VN065379.D
Acq On : 5 Jan 2021 11:11
Operator : JC/MD
Sample : VN0105MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0105MBL01

Quant Time: Jan 06 01:14:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 22 15:42:18 2020
Response via : Initial Calibration

