

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010521\
 Data File : VN065383.D
 Acq On : 5 Jan 2021 13:08
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Jan 06 01:15:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.630	168	277466	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.553	114	469446	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.380	117	492158	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.318	152	189971	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.991	65	156587	49.99	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.98%
35) Dibromofluoromethane	7.553	113	143353	50.62	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.24%
50) Toluene-d8	10.058	98	582155	54.84	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	109.68%
62) 4-Bromofluorobenzene	12.376	95	218102	51.75	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	103.50%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010521\
Data File : VN065383.D
Acq On : 5 Jan 2021 13:08
Operator : JC/MD
Sample : IBLK
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Jan 06 01:15:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 22 15:42:18 2020
Response via : Initial Calibration

