

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010521\
 Data File : VN065385.D
 Acq On : 5 Jan 2021 14:23
 Operator : JC/MD
 Sample : L5248-05
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VR16125

Quant Time: Jan 06 01:16:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.631	168	261903	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.553	114	442099	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.380	117	472417	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.318	152	188748	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.991	65	150744	50.98	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 101.96%			
35) Dibromofluoromethane	7.557	113	135625	50.85	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 101.70%			
50) Toluene-d8	10.061	98	554011	55.42	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 110.84%			
62) 4-Bromofluorobenzene	12.376	95	209014	52.58	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 105.16%			
Target Compounds						Qvalue
16) Acetone	3.795	43	31488	15.28	ug/l	97
20) Methylene Chloride	4.512	84	6166	2.14	ug/l	98
29) Tetrahydrofuran	7.184	42	10005	9.32	ug/l	96
52) Toluene	10.126	92	23538	3.12	ug/l	98
88) 1,4-Dichlorobenzene	13.338	146	61238	9.71	ug/l	98
95) Naphthalene	15.113	128	3632	3.72	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010521\
Data File : VN065385.D
Acq On : 5 Jan 2021 14:23
Operator : JC/MD
Sample : L5248-05
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VR16125

Quant Time: Jan 06 01:16:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 22 15:42:18 2020
Response via : Initial Calibration

